



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

ELAP Certificates 1551. 2728. 2922. and 2303

28 December 2018

Sample Traps, LLC

Attn: Quality Control Manager

262 Rickenbacker Circle

Livermore, CA 94551

RE: QC- 40ml Amber VOA (NP)

Work Order: 18L1885

Enclosed are the results of analyses for samples received by the laboratory on 12/17/18 07:30. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Jeanette L. Poplin For Chelsea L. Sandelin
Project Manager



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
A8337CUBS - 01	18L1885-01	Water	12/17/18 00:00	12/17/18 07:30
A8337CUBS - 02	18L1885-02	Water	12/17/18 00:00	12/17/18 07:30



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
A8337CUBS - 02 (18L1885-02) Water Sampled: 12/17/18 00:00 Received: 12/17/18 07:30												
Acetone	ND	2.0	5.0		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Acrylonitrile	ND	0.40	5.0		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Benzene	ND	0.30	0.30		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Bromobenzene	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Bromochloromethane	ND	0.40	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Bromodichloromethane	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Bromoform	ND	0.30	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Bromomethane	ND	0.40	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
n-Butylbenzene	ND	0.50	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
sec-Butylbenzene	ND	0.50	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
tert-Butylbenzene	ND	0.50	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Carbon disulfide	ND	0.40	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Carbon tetrachloride	ND	0.30	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Chlorobenzene	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Chloroethane	ND	0.30	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Chloroform	ND	0.30	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Chloromethane	ND	0.40	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
2-Chlorotoluene	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
4-Chlorotoluene	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Dibromochloromethane	ND	0.30	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,2-Dibromo-3-chloropropane	ND	0.50	1.0		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,2-Dibromoethane (EDB)	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Dibromomethane	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,2-Dichlorobenzene	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,3-Dichlorobenzene	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,4-Dichlorobenzene	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
trans-1,4-Dichloro-2-butene	ND	0.90	5.0		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Dichlorodifluoromethane	ND	0.50	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,1-Dichloroethane	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,2-Dichloroethane	ND	0.50	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,1-Dichloroethene	ND	0.30	0.30		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
cis-1,2-Dichloroethene	ND	0.40	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
trans-1,2-Dichloroethene	ND	0.40	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
1,2-Dichloropropane	ND	0.20	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
A8337CUBS - 02 (18L1885-02) Water Sampled: 12/17/18 00:00 Received: 12/17/18 07:30												
1,3-Dichloropropane	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
2,2-Dichloropropane	ND	0.30	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,1-Dichloropropene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
cis-1,3-Dichloropropene	ND	0.30	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
trans-1,3-Dichloropropene	ND	0.30	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,3-Dichloropropene (total)	ND	0.30	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
2-Hexanone	ND	0.50	5.0	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Ethylbenzene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Hexachlorobutadiene	ND	0.40	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Isopropylbenzene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
p-Isopropyltoluene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Methyl ethyl ketone	ND	0.60	1.0	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Methyl iodide	ND	0.40	2.0	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Methyl isobutyl ketone	ND	0.80	1.0	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Methylene chloride	ND	0.40	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Naphthalene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
n-Propylbenzene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Styrene	ND	0.20	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,1,1,2-Tetrachloroethane	ND	0.40	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,1,2,2-Tetrachloroethane	ND	0.20	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Tetrachloroethene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Toluene	ND	0.30	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,2,3-Trichlorobenzene	ND	0.40	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,2,4-Trichlorobenzene	ND	0.40	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,1,1-Trichloroethane	ND	0.40	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,1,2-Trichloroethane	ND	0.20	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Trichloroethene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Trichlorofluoromethane	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Trichlorotrifluoroethane	ND	0.40	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,2,3-Trichloropropane	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,2,4-Trimethylbenzene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
1,3,5-Trimethylbenzene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
Vinyl chloride	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U
m,p-Xylene	ND	0.50	0.50	ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV		U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
A8337CUBS - 02 (18L1885-02) Water Sampled: 12/17/18 00:00 Received: 12/17/18 07:30												
o-Xylene	ND	0.50	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Xylenes (total)	ND	0.50	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Trihalomethanes (total)	ND	0.30	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Methyl tert-butyl ether	ND	0.50	3.0		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Ethyl tert-butyl ether	ND	0.40	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Tert-amyl methyl ether	ND	0.30	0.50		ug/L	1	AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	U
Surrogate: Bromofluorobenzene		98.8 %	70-130				AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	
Surrogate: Dibromofluoromethane		92.4 %	70-130				AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	
Surrogate: Toluene-d8		109 %	70-130				AL84158	12/20/18 10:00	12/20/18 13:22	EPA 524.2	JV	



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
A8337CUBS - 01 (18L1885-01) Water Sampled: 12/17/18 00:00 Received: 12/17/18 07:30												
Acetone	ND	3.0	5.0	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Acetonitrile	ND	50	100	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Acrylonitrile	ND	0.40	5.0	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Allyl chloride	ND	0.40	10	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Benzene	ND	0.30	0.30	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Bromobenzene	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Bromochloromethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Bromodichloromethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Bromoform	ND	0.30	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Bromomethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
n-Butylbenzene	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
sec-Butylbenzene	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
tert-Butylbenzene	ND	0.30	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Carbon disulfide	ND	0.40	5.0	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Carbon tetrachloride	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Chlorobenzene	ND	0.30	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Chloroethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Chloroform	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Chloromethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Chloroprene	ND	0.40	1.0	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
2-Chlorotoluene	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
4-Chlorotoluene	ND	0.30	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Dibromochloromethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,2-Dibromo-3-chloropropane	ND	0.60	2.0	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,2-Dibromoethane (EDB)	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Dibromomethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,2-Dichlorobenzene	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,3-Dichlorobenzene	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,4-Dichlorobenzene	ND	0.10	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
trans-1,4-Dichloro-2-butene	ND	0.50	5.0	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
Dichlorodifluoromethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,1-Dichloroethane	ND	0.10	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,2-Dichloroethane	ND	0.40	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	
1,1-Dichloroethene	ND	0.30	0.50	ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
A8337CUBS - 01 (18L1885-01) Water Sampled: 12/17/18 00:00 Received: 12/17/18 07:30												
cis-1,2-Dichloroethene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
trans-1,2-Dichloroethene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,2-Dichloropropane	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,3-Dichloropropane	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
2,2-Dichloropropane	ND	0.20	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,1-Dichloropropene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
cis-1,3-Dichloropropene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
trans-1,3-Dichloropropene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Diethyl ether	ND	0.20	1.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Di-isopropyl ether	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Ethanol	ND	50	50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Ethyl methacrylate	ND	0.70	10		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Ethylbenzene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Ethyl tert-butyl ether	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Hexachlorobutadiene	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Hexachloroethane	ND	0.40	1.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
2-Hexanone	ND	0.50	5.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Isobutanol	ND	40	100		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Isopropylbenzene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
p-Isopropyltoluene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Methacrylonitrile	ND	0.40	1.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Methylene chloride	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Methyl ethyl ketone	ND	0.70	1.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Methyl iodide	ND	0.40	2.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Methyl isobutyl ketone	ND	0.60	1.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Methyl methacrylate	ND	0.40	1.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Methyl tert-butyl ether	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Naphthalene	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Propionitrile	ND	20	50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
n-Propylbenzene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Styrene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Tert-amyl methyl ether	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Tert-butyl alcohol	ND	6.0	10		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,1,1,2-Tetrachloroethane	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
A8337CUBS - 01 (18L1885-01) Water Sampled: 12/17/18 00:00 Received: 12/17/18 07:30												
1,1,2,2-Tetrachloroethane	ND	0.30	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Tetrachloroethene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Tetrahydrofuran	ND	0.40	5.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Toluene	ND	0.30	0.30		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,2,3-Trichlorobenzene	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,2,4-Trichlorobenzene	ND	0.20	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,1,1-Trichloroethane	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,1,2-Trichloroethane	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Trichloroethene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Trichlorofluoromethane	ND	0.20	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,2,3-Trichloropropane	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Trichlorotrifluoroethane	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,2,4-Trimethylbenzene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
1,3,5-Trimethylbenzene	ND	0.30	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Vinyl acetate	ND	0.80	1.0		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Vinyl chloride	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
m,p-Xylene	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
o-Xylene	ND	0.40	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
Xylenes (total)	ND	0.50	0.50		ug/L	1	AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	U
<i>Surrogate: Bromofluorobenzene</i>			107 %				AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	
<i>Surrogate: Dibromofluoromethane</i>			99.6 %				AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	
<i>Surrogate: Toluene-d8</i>			113 %				AL83983	12/18/18 10:00	12/18/18 11:53	EPA 8260	JV	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

Blank (AL84158-BLK1)

Prepared: 12/19/18 Analyzed: 12/20/18

Acetone	ND	2.0	5.0	ug/L							U
Acrylonitrile	ND	0.40	5.0	ug/L							U
Benzene	ND	0.30	0.30	ug/L							U
Bromobenzene	ND	0.20	0.50	ug/L							U
Bromochloromethane	ND	0.40	0.50	ug/L							U
Bromodichloromethane	ND	0.20	0.50	ug/L							U
Bromoform	ND	0.30	0.50	ug/L							U
Bromomethane	ND	0.40	0.50	ug/L							U
n-Butylbenzene	ND	0.50	0.50	ug/L							U
sec-Butylbenzene	ND	0.50	0.50	ug/L							U
tert-Butylbenzene	ND	0.50	0.50	ug/L							U
Carbon disulfide	ND	0.40	0.50	ug/L							U
Carbon tetrachloride	ND	0.30	0.50	ug/L							U
Chlorobenzene	ND	0.20	0.50	ug/L							U
Chloroethane	ND	0.30	0.50	ug/L							U
Chloroform	ND	0.30	0.50	ug/L							U
Chloromethane	ND	0.40	0.50	ug/L							U
2-Chlorotoluene	ND	0.20	0.50	ug/L							U
4-Chlorotoluene	ND	0.20	0.50	ug/L							U
Dibromochloromethane	ND	0.30	0.50	ug/L							U
1,2-Dibromo-3-chloropropane	ND	0.50	1.0	ug/L							U
1,2-Dibromoethane (EDB)	ND	0.20	0.50	ug/L							U
Dibromomethane	ND	0.20	0.50	ug/L							U
1,2-Dichlorobenzene	ND	0.20	0.50	ug/L							U
1,3-Dichlorobenzene	ND	0.20	0.50	ug/L							U
trans-1,4-Dichloro-2-butene	ND	0.90	5.0	ug/L							U
1,4-Dichlorobenzene	ND	0.20	0.50	ug/L							U
Dichlorodifluoromethane	ND	0.50	0.50	ug/L							U
1,1-Dichloroethane	ND	0.20	0.50	ug/L							U
1,2-Dichloroethane	ND	0.50	0.50	ug/L							U
1,1-Dichloroethene	ND	0.30	0.30	ug/L							U
cis-1,2-Dichloroethene	ND	0.40	0.50	ug/L							U
trans-1,2-Dichloroethene	ND	0.40	0.50	ug/L							U
1,2-Dichloropropane	ND	0.20	0.50	ug/L							U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

Blank (AL84158-BLK1)

Prepared: 12/19/18 Analyzed: 12/20/18

1,3-Dichloropropane	ND	0.50	0.50	ug/L							U
2,2-Dichloropropane	ND	0.30	0.50	ug/L							U
1,1-Dichloropropene	ND	0.50	0.50	ug/L							U
cis-1,3-Dichloropropene	ND	0.30	0.50	ug/L							U
trans-1,3-Dichloropropene	ND	0.30	0.50	ug/L							U
2-Hexanone	ND	0.50	5.0	ug/L							U
1,3-Dichloropropene (total)	ND	0.30	0.50	ug/L							U
Ethylbenzene	ND	0.50	0.50	ug/L							U
Hexachlorobutadiene	ND	0.40	0.50	ug/L							U
Isopropylbenzene	ND	0.50	0.50	ug/L							U
p-Isopropyltoluene	ND	0.50	0.50	ug/L							U
Methyl ethyl ketone	ND	0.60	1.0	ug/L							U
Methyl iodide	ND	0.40	2.0	ug/L							U
Methyl isobutyl ketone	ND	0.80	1.0	ug/L							U
Methylene chloride	ND	0.40	0.50	ug/L							U
Naphthalene	ND	0.50	0.50	ug/L							U
n-Propylbenzene	ND	0.50	0.50	ug/L							U
Styrene	ND	0.20	0.50	ug/L							U
1,1,1,2-Tetrachloroethane	ND	0.40	0.50	ug/L							U
1,1,2,2-Tetrachloroethane	ND	0.20	0.50	ug/L							U
Tetrachloroethene	ND	0.50	0.50	ug/L							U
Toluene	ND	0.30	0.50	ug/L							U
1,2,3-Trichlorobenzene	ND	0.40	0.50	ug/L							U
1,2,4-Trichlorobenzene	ND	0.40	0.50	ug/L							U
1,1,1-Trichloroethane	ND	0.40	0.50	ug/L							U
1,1,2-Trichloroethane	ND	0.20	0.50	ug/L							U
Trichloroethene	ND	0.50	0.50	ug/L							U
Trichlorofluoromethane	ND	0.50	0.50	ug/L							U
Trichlorotrifluoroethane	ND	0.40	0.50	ug/L							U
1,2,3-Trichloropropane	ND	0.50	0.50	ug/L							U
1,2,4-Trimethylbenzene	ND	0.50	0.50	ug/L							U
1,3,5-Trimethylbenzene	ND	0.50	0.50	ug/L							U
Vinyl chloride	ND	0.50	0.50	ug/L							U
m,p-Xylene	ND	0.50	0.50	ug/L							U
o-Xylene	ND	0.50	0.50	ug/L							U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting		Spike Level	Source		%REC		RPD	Limit	Notes
			Limit	Units		Result	%REC	Limits	RPD			

Batch AL84158 - VOAs in Water GCMS

Blank (AL84158-BLK1)

Prepared: 12/19/18 Analyzed: 12/20/18

Xylenes (total)	ND	0.50	0.50	ug/L								U
Trihalomethanes (total)	ND	0.30	0.50	ug/L								U
Methyl tert-butyl ether	ND	0.50	3.0	ug/L								U
Ethyl tert-butyl ether	ND	0.40	0.50	ug/L								U
Tert-amyl methyl ether	ND	0.30	0.50	ug/L								U
Surrogate: Bromofluorobenzene	25.4			ug/L	25.0		102	70-130				
Surrogate: Dibromofluoromethane	24.0			ug/L	25.0		95.8	70-130				
Surrogate: Toluene-d8	27.9			ug/L	25.0		112	70-130				

LCS (AL84158-BS1)

Prepared: 12/19/18 Analyzed: 12/20/18

Acetone	19.0	2.0	5.0	ug/L	20.0		95.1	70-130				
Acrylonitrile	5.21	0.40	5.0	ug/L	5.00		104	70-130				
Benzene	4.82	0.30	0.30	ug/L	5.00		96.4	70-130				
Bromobenzene	4.76	0.20	0.50	ug/L	5.00		95.2	70-130				
Bromochloromethane	4.67	0.40	0.50	ug/L	5.00		93.4	70-130				
Bromodichloromethane	5.96	0.20	0.50	ug/L	5.00		119	70-130				
Bromoform	4.78	0.30	0.50	ug/L	5.00		95.6	70-130				
Bromomethane	4.73	0.40	0.50	ug/L	5.00		94.6	70-130				
n-Butylbenzene	4.59	0.50	0.50	ug/L	5.00		91.8	70-130				
sec-Butylbenzene	4.87	0.50	0.50	ug/L	5.00		97.4	70-130				
tert-Butylbenzene	4.93	0.50	0.50	ug/L	5.00		98.6	70-130				
Carbon disulfide	3.97	0.40	0.50	ug/L	5.00		79.4	70-130				
Carbon tetrachloride	5.00	0.30	0.50	ug/L	5.00		100	70-130				
Chlorobenzene	4.74	0.20	0.50	ug/L	5.00		94.8	70-130				
Chloroethane	4.64	0.30	0.50	ug/L	5.00		92.8	70-130				
Chloroform	4.90	0.30	0.50	ug/L	5.00		98.0	70-130				
Chloromethane	5.45	0.40	0.50	ug/L	5.00		109	70-130				
2-Chlorotoluene	4.93	0.20	0.50	ug/L	5.00		98.6	70-130				
4-Chlorotoluene	4.91	0.20	0.50	ug/L	5.00		98.2	70-130				
Dibromochloromethane	5.79	0.30	0.50	ug/L	5.00		116	70-130				
1,2-Dibromo-3-chloropropane	4.63	0.50	1.0	ug/L	5.00		92.6	70-130				
1,2-Dibromoethane (EDB)	5.07	0.20	0.50	ug/L	5.00		101	70-130				
Dibromomethane	4.37	0.20	0.50	ug/L	5.00		87.4	70-130				
1,2-Dichlorobenzene	4.63	0.20	0.50	ug/L	5.00		92.6	70-130				
1,3-Dichlorobenzene	4.76	0.20	0.50	ug/L	5.00		95.2	70-130				

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting		Spike Level	Source		%REC		RPD		Notes
			Limit	Units		Result	%REC	Limits	RPD	Limit		

Batch AL84158 - VOAs in Water GCMS

LCS (AL84158-BS1)

Prepared: 12/19/18 Analyzed: 12/20/18

trans-1,4-Dichloro-2-butene	3.92	0.90	5.0	ug/L	5.00	78.4	70-130					J
1,4-Dichlorobenzene	4.57	0.20	0.50	ug/L	5.00	91.4	70-130					
Dichlorodifluoromethane	5.72	0.50	0.50	ug/L	5.00	114	70-130					
1,1-Dichloroethane	4.91	0.20	0.50	ug/L	5.00	98.2	70-130					
1,2-Dichloroethane	4.50	0.50	0.50	ug/L	5.00	90.0	70-130					
1,1-Dichloroethene	4.41	0.30	0.30	ug/L	5.00	88.2	70-130					
cis-1,2-Dichloroethene	4.63	0.40	0.50	ug/L	5.00	92.6	70-130					
trans-1,2-Dichloroethene	4.70	0.40	0.50	ug/L	5.00	94.0	70-130					
1,2-Dichloropropane	4.57	0.20	0.50	ug/L	5.00	91.4	70-130					
1,3-Dichloropropane	4.85	0.50	0.50	ug/L	5.00	97.0	70-130					
2,2-Dichloropropane	3.67	0.30	0.50	ug/L	5.00	73.4	70-130					
1,1-Dichloropropene	4.89	0.50	0.50	ug/L	5.00	97.8	70-130					
cis-1,3-Dichloropropene	4.03	0.30	0.50	ug/L	5.00	80.6	70-130					
trans-1,3-Dichloropropene	4.04	0.30	0.50	ug/L	5.00	80.8	70-130					
Ethylbenzene	4.86	0.50	0.50	ug/L	5.00	97.2	70-130					
2-Hexanone	5.44	0.50	5.0	ug/L	5.00	109	70-130					
Hexachlorobutadiene	3.80	0.40	0.50	ug/L	5.00	76.0	70-130					
Isopropylbenzene	5.20	0.50	0.50	ug/L	5.00	104	70-130					
p-Isopropyltoluene	4.90	0.50	0.50	ug/L	5.00	98.0	70-130					
Methyl ethyl ketone	10.8	0.60	1.0	ug/L	10.0	108	70-130					
Methyl iodide	4.76	0.40	2.0	ug/L	5.00	95.2	70-130					
Methyl isobutyl ketone	10.7	0.80	1.0	ug/L	10.0	107	70-130					
Methylene chloride	3.61	0.40	0.50	ug/L	5.00	72.2	70-130					
Naphthalene	4.98	0.50	0.50	ug/L	5.00	99.6	70-130					
n-Propylbenzene	4.86	0.50	0.50	ug/L	5.00	97.2	70-130					
Styrene	4.84	0.20	0.50	ug/L	5.00	96.8	70-130					
1,1,1,2-Tetrachloroethane	4.38	0.40	0.50	ug/L	5.00	87.6	70-130					
1,1,2,2-Tetrachloroethane	4.92	0.20	0.50	ug/L	5.00	98.4	70-130					
Tetrachloroethene	4.45	0.50	0.50	ug/L	5.00	89.0	70-130					
Toluene	5.09	0.30	0.50	ug/L	5.00	102	70-130					
1,2,3-Trichlorobenzene	4.54	0.40	0.50	ug/L	5.00	90.8	70-130					
1,2,4-Trichlorobenzene	4.31	0.40	0.50	ug/L	5.00	86.2	70-130					
1,1,1-Trichloroethane	5.06	0.40	0.50	ug/L	5.00	101	70-130					
1,1,2-Trichloroethane	4.79	0.20	0.50	ug/L	5.00	95.8	70-130					

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

LCS (AL84158-BS1)

Prepared: 12/19/18 Analyzed: 12/20/18

Trichloroethene	4.81	0.50	0.50	ug/L	5.00		96.2	70-130			
Trichlorofluoromethane	5.75	0.50	0.50	ug/L	5.00		115	70-130			
Trichlorotrifluoroethane	5.04	0.40	0.50	ug/L	5.00		101	70-130			
1,2,3-Trichloropropane	5.04	0.50	0.50	ug/L	5.00		101	70-130			
1,2,4-Trimethylbenzene	5.06	0.50	0.50	ug/L	5.00		101	70-130			
1,3,5-Trimethylbenzene	4.89	0.50	0.50	ug/L	5.00		97.8	70-130			
Vinyl chloride	5.20	0.50	0.50	ug/L	5.00		104	70-130			
m,p-Xylene	9.71	0.50	0.50	ug/L	10.0		97.1	70-130			
o-Xylene	4.90	0.50	0.50	ug/L	5.00		98.0	70-130			
Xylenes (total)	14.6	0.50	0.50	ug/L	15.0		97.4	70-130			
Methyl tert-butyl ether	4.17	0.50	3.0	ug/L	5.00		83.4	70-130			
Ethyl tert-butyl ether	5.16	0.40	0.50	ug/L	5.00		103	70-130			
Tert-amyl methyl ether	4.07	0.30	0.50	ug/L	5.00		81.4	70-130			
Surrogate: Bromofluorobenzene	25.4			ug/L	25.0		101	70-130			
Surrogate: Dibromofluoromethane	23.7			ug/L	25.0		94.7	70-130			
Surrogate: Toluene-d8	26.2			ug/L	25.0		105	70-130			

LCS Dup (AL84158-BSD1)

Prepared: 12/19/18 Analyzed: 12/20/18

Acetone	18.2	2.0	5.0	ug/L	20.0		90.8	70-130	4.68	30	
Acrylonitrile	5.02	0.40	5.0	ug/L	5.00		100	70-130	3.71	30	
Benzene	4.80	0.30	0.30	ug/L	5.00		96.0	70-130	0.416	30	
Bromobenzene	4.81	0.20	0.50	ug/L	5.00		96.2	70-130	1.04	30	
Bromochloromethane	4.69	0.40	0.50	ug/L	5.00		93.8	70-130	0.427	30	
Bromodichloromethane	6.12	0.20	0.50	ug/L	5.00		122	70-130	2.65	30	
Bromoform	4.73	0.30	0.50	ug/L	5.00		94.6	70-130	1.05	30	
Bromomethane	4.82	0.40	0.50	ug/L	5.00		96.4	70-130	1.88	30	
n-Butylbenzene	4.54	0.50	0.50	ug/L	5.00		90.8	70-130	1.10	30	
sec-Butylbenzene	4.79	0.50	0.50	ug/L	5.00		95.8	70-130	1.66	30	
tert-Butylbenzene	4.94	0.50	0.50	ug/L	5.00		98.8	70-130	0.203	30	
Carbon disulfide	4.24	0.40	0.50	ug/L	5.00		84.8	70-130	6.58	30	
Carbon tetrachloride	5.19	0.30	0.50	ug/L	5.00		104	70-130	3.73	30	
Chlorobenzene	4.63	0.20	0.50	ug/L	5.00		92.6	70-130	2.35	30	
Chloroethane	4.61	0.30	0.50	ug/L	5.00		92.2	70-130	0.649	30	
Chloroform	4.99	0.30	0.50	ug/L	5.00		99.8	70-130	1.82	30	
Chloromethane	5.97	0.40	0.50	ug/L	5.00		119	70-130	9.11	30	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

LCS Dup (AL84158-BSD1)

Prepared: 12/19/18 Analyzed: 12/20/18

2-Chlorotoluene	4.92	0.20	0.50	ug/L	5.00		98.4	70-130	0.203	30	
4-Chlorotoluene	4.79	0.20	0.50	ug/L	5.00		95.8	70-130	2.47	30	
Dibromochloromethane	5.85	0.30	0.50	ug/L	5.00		117	70-130	1.03	30	
1,2-Dibromo-3-chloropropane	4.76	0.50	1.0	ug/L	5.00		95.2	70-130	2.77	25	
1,2-Dibromoethane (EDB)	5.02	0.20	0.50	ug/L	5.00		100	70-130	0.991	25	
Dibromomethane	4.40	0.20	0.50	ug/L	5.00		88.0	70-130	0.684	30	
1,2-Dichlorobenzene	4.64	0.20	0.50	ug/L	5.00		92.8	70-130	0.216	30	
1,3-Dichlorobenzene	4.62	0.20	0.50	ug/L	5.00		92.4	70-130	2.99	30	
trans-1,4-Dichloro-2-butene	4.00	0.90	5.0	ug/L	5.00		80.0	70-130	2.02	25	J
1,4-Dichlorobenzene	4.61	0.20	0.50	ug/L	5.00		92.2	70-130	0.871	30	
Dichlorodifluoromethane	5.97	0.50	0.50	ug/L	5.00		119	70-130	4.28	30	
1,1-Dichloroethane	4.95	0.20	0.50	ug/L	5.00		99.0	70-130	0.811	30	
1,2-Dichloroethane	4.39	0.50	0.50	ug/L	5.00		87.8	70-130	2.47	30	
1,1-Dichloroethene	4.56	0.30	0.30	ug/L	5.00		91.2	70-130	3.34	30	
cis-1,2-Dichloroethene	4.72	0.40	0.50	ug/L	5.00		94.4	70-130	1.93	30	
trans-1,2-Dichloroethene	4.76	0.40	0.50	ug/L	5.00		95.2	70-130	1.27	30	
1,2-Dichloropropane	4.61	0.20	0.50	ug/L	5.00		92.2	70-130	0.871	30	
1,3-Dichloropropane	4.79	0.50	0.50	ug/L	5.00		95.8	70-130	1.24	30	
2,2-Dichloropropane	3.65	0.30	0.50	ug/L	5.00		73.0	70-130	0.546	30	
1,1-Dichloropropene	4.99	0.50	0.50	ug/L	5.00		99.8	70-130	2.02	30	
cis-1,3-Dichloropropene	4.06	0.30	0.50	ug/L	5.00		81.2	70-130	0.742	30	
trans-1,3-Dichloropropene	4.12	0.30	0.50	ug/L	5.00		82.4	70-130	1.96	30	
2-Hexanone	5.55	0.50	5.0	ug/L	5.00		111	70-130	2.00	25	
Ethylbenzene	4.68	0.50	0.50	ug/L	5.00		93.6	70-130	3.77	30	
Hexachlorobutadiene	3.51	0.40	0.50	ug/L	5.00		70.2	70-130	7.93	30	
Isopropylbenzene	5.16	0.50	0.50	ug/L	5.00		103	70-130	0.772	30	
p-Isopropyltoluene	4.77	0.50	0.50	ug/L	5.00		95.4	70-130	2.69	30	
Methyl ethyl ketone	10.2	0.60	1.0	ug/L	10.0		102	70-130	6.20	30	
Methyl iodide	4.79	0.40	2.0	ug/L	5.00		95.8	70-130	0.628	25	
Methyl isobutyl ketone	11.2	0.80	1.0	ug/L	10.0		112	70-130	4.46	30	
Methylene chloride	3.61	0.40	0.50	ug/L	5.00		72.2	70-130	0.00	30	
Naphthalene	4.77	0.50	0.50	ug/L	5.00		95.4	70-130	4.31	30	
n-Propylbenzene	4.91	0.50	0.50	ug/L	5.00		98.2	70-130	1.02	30	
Styrene	4.64	0.20	0.50	ug/L	5.00		92.8	70-130	4.22	30	
1,1,1,2-Tetrachloroethane	4.37	0.40	0.50	ug/L	5.00		87.4	70-130	0.229	30	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

LCS Dup (AL84158-BSD1)

Prepared: 12/19/18 Analyzed: 12/20/18

1,1,2,2-Tetrachloroethane	4.91	0.20	0.50	ug/L	5.00		98.2	70-130	0.203	30	
Tetrachloroethene	4.57	0.50	0.50	ug/L	5.00		91.4	70-130	2.66	30	
Toluene	5.03	0.30	0.50	ug/L	5.00		101	70-130	1.19	30	
1,2,3-Trichlorobenzene	4.35	0.40	0.50	ug/L	5.00		87.0	70-130	4.27	30	
1,2,4-Trichlorobenzene	4.06	0.40	0.50	ug/L	5.00		81.2	70-130	5.97	30	
1,1,1-Trichloroethane	5.30	0.40	0.50	ug/L	5.00		106	70-130	4.63	30	
1,1,1,2-Trichloroethane	4.76	0.20	0.50	ug/L	5.00		95.2	70-130	0.628	30	
Trichloroethene	4.83	0.50	0.50	ug/L	5.00		96.6	70-130	0.415	30	
Trichlorofluoromethane	6.18	0.50	0.50	ug/L	5.00		124	70-130	7.21	30	
Trichlorotrifluoroethane	5.19	0.40	0.50	ug/L	5.00		104	70-130	2.93	30	
1,2,3-Trichloropropane	5.00	0.50	0.50	ug/L	5.00		100	70-130	0.797	25	
1,2,4-Trimethylbenzene	4.93	0.50	0.50	ug/L	5.00		98.6	70-130	2.60	30	
1,3,5-Trimethylbenzene	4.94	0.50	0.50	ug/L	5.00		98.8	70-130	1.02	30	
Vinyl chloride	5.15	0.50	0.50	ug/L	5.00		103	70-130	0.966	30	
m,p-Xylene	9.40	0.50	0.50	ug/L	10.0		94.0	70-130	3.24	30	
o-Xylene	4.69	0.50	0.50	ug/L	5.00		93.8	70-130	4.38	30	
Xylenes (total)	14.1	0.50	0.50	ug/L	15.0		93.9	70-130	3.62	30	
Methyl tert-butyl ether	4.07	0.50	3.0	ug/L	5.00		81.4	70-130	2.43	30	
Ethyl tert-butyl ether	5.17	0.40	0.50	ug/L	5.00		103	70-130	0.194	30	
Tert-amyl methyl ether	4.11	0.30	0.50	ug/L	5.00		82.2	70-130	0.978	30	
Surrogate: Bromofluorobenzene	26.8			ug/L	25.0		107	70-130			
Surrogate: Dibromofluoromethane	24.8			ug/L	25.0		99.0	70-130			
Surrogate: Toluene-d8	27.0			ug/L	25.0		108	70-130			

Matrix Spike (AL84158-MS1)

Source: 18L1288-01

Prepared: 12/19/18 Analyzed: 12/20/18

Acetone	27.9	2.0	5.0	ug/L	20.0	4.77	116	70-130			
Acrylonitrile	5.91	0.40	5.0	ug/L	5.00	ND	118	70-130			
Benzene	5.22	0.30	0.30	ug/L	5.00	ND	104	70-130			
Bromobenzene	4.94	0.20	0.50	ug/L	5.00	ND	98.8	70-130			
Bromochloromethane	4.85	0.40	0.50	ug/L	5.00	ND	97.0	70-130			
Bromodichloromethane	6.45	0.20	0.50	ug/L	5.00	ND	129	70-130			
Bromoform	4.79	0.30	0.50	ug/L	5.00	ND	95.8	70-130			
Bromomethane	5.47	0.40	0.50	ug/L	5.00	ND	109	70-130			
n-Butylbenzene	5.41	0.50	0.50	ug/L	5.00	ND	108	70-130			
sec-Butylbenzene	5.54	0.50	0.50	ug/L	5.00	ND	111	70-130			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

Matrix Spike (AL84158-MS1)

Source: 18L1288-01

Prepared: 12/19/18 Analyzed: 12/20/18

tert-Butylbenzene	5.49	0.50	0.50	ug/L	5.00	ND	110	70-130			
Carbon disulfide	4.53	0.40	0.50	ug/L	5.00	ND	90.6	70-130			
Carbon tetrachloride	5.73	0.30	0.50	ug/L	5.00	ND	115	70-130			
Chlorobenzene	4.94	0.20	0.50	ug/L	5.00	ND	98.8	70-130			
Chloroethane	5.40	0.30	0.50	ug/L	5.00	ND	108	70-130			
Chloroform	5.36	0.30	0.50	ug/L	5.00	ND	107	70-130			
Chloromethane	4.48	0.40	0.50	ug/L	5.00	ND	89.6	70-130			
2-Chlorotoluene	5.34	0.20	0.50	ug/L	5.00	ND	107	70-130			
4-Chlorotoluene	5.17	0.20	0.50	ug/L	5.00	ND	103	70-130			
Dibromochloromethane	5.78	0.30	0.50	ug/L	5.00	ND	116	70-130			
1,2-Dibromo-3-chloropropane	4.43	0.50	1.0	ug/L	5.00	ND	88.6	70-130			
1,2-Dibromoethane (EDB)	5.21	0.20	0.50	ug/L	5.00	ND	104	70-130			
Dibromomethane	4.62	0.20	0.50	ug/L	5.00	ND	92.4	70-130			
1,2-Dichlorobenzene	4.76	0.20	0.50	ug/L	5.00	ND	95.2	70-130			
1,3-Dichlorobenzene	4.99	0.20	0.50	ug/L	5.00	ND	99.8	70-130			
1,4-Dichlorobenzene	4.71	0.20	0.50	ug/L	5.00	ND	94.2	70-130			
trans-1,4-Dichloro-2-butene	4.61	0.90	5.0	ug/L	5.00	ND	92.2	70-130			J
Dichlorodifluoromethane	6.70	0.50	0.50	ug/L	5.00	ND	134	70-130			QM-05
1,1-Dichloroethane	5.25	0.20	0.50	ug/L	5.00	ND	105	70-130			
1,2-Dichloroethane	4.63	0.50	0.50	ug/L	5.00	ND	92.6	70-130			
1,1-Dichloroethene	4.85	0.30	0.30	ug/L	5.00	ND	97.0	70-130			
cis-1,2-Dichloroethene	4.95	0.40	0.50	ug/L	5.00	ND	99.0	70-130			
trans-1,2-Dichloroethene	5.09	0.40	0.50	ug/L	5.00	ND	102	70-130			
1,2-Dichloropropane	4.95	0.20	0.50	ug/L	5.00	ND	99.0	70-130			
1,3-Dichloropropane	4.95	0.50	0.50	ug/L	5.00	ND	99.0	70-130			
2,2-Dichloropropane	5.46	0.30	0.50	ug/L	5.00	ND	109	70-130			
1,1-Dichloropropene	5.35	0.50	0.50	ug/L	5.00	ND	107	70-130			
cis-1,3-Dichloropropene	4.32	0.30	0.50	ug/L	5.00	ND	86.4	70-130			
trans-1,3-Dichloropropene	4.33	0.30	0.50	ug/L	5.00	ND	86.6	70-130			
2-Hexanone	5.75	0.50	5.0	ug/L	5.00	ND	115	70-130			
Ethylbenzene	5.23	0.50	0.50	ug/L	5.00	ND	105	70-130			
Hexachlorobutadiene	4.48	0.40	0.50	ug/L	5.00	ND	89.6	70-130			
Isopropylbenzene	5.69	0.50	0.50	ug/L	5.00	ND	114	70-130			
p-Isopropyltoluene	5.50	0.50	0.50	ug/L	5.00	ND	110	70-130			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

Matrix Spike (AL84158-MS1)	Source: 18L1288-01	Prepared: 12/19/18	Analyzed: 12/20/18					
Methyl ethyl ketone	12.3	0.60	1.0 ug/L	10.0	ND	123	70-130	
Methyl iodide	5.08	0.40	2.0 ug/L	5.00	ND	102	70-130	
Methyl isobutyl ketone	11.8	0.80	1.0 ug/L	10.0	ND	118	70-130	
Methylene chloride	3.58	0.40	0.50 ug/L	5.00	ND	71.6	70-130	
Naphthalene	4.88	0.50	0.50 ug/L	5.00	ND	97.6	70-130	
n-Propylbenzene	5.42	0.50	0.50 ug/L	5.00	ND	108	70-130	
Styrene	5.09	0.20	0.50 ug/L	5.00	ND	102	70-130	
1,1,1,2-Tetrachloroethane	4.70	0.40	0.50 ug/L	5.00	ND	94.0	70-130	
1,1,1,2,2-Tetrachloroethane	5.04	0.20	0.50 ug/L	5.00	ND	101	70-130	
Tetrachloroethene	4.94	0.50	0.50 ug/L	5.00	ND	98.8	70-130	
Toluene	5.49	0.30	0.50 ug/L	5.00	ND	110	70-130	
1,2,3-Trichlorobenzene	4.49	0.40	0.50 ug/L	5.00	ND	89.8	70-130	
1,2,4-Trichlorobenzene	4.48	0.40	0.50 ug/L	5.00	ND	89.6	70-130	
1,1,1-Trichloroethane	5.67	0.40	0.50 ug/L	5.00	ND	113	70-130	
1,1,2-Trichloroethane	5.02	0.20	0.50 ug/L	5.00	ND	100	70-130	
Trichloroethene	5.13	0.50	0.50 ug/L	5.00	ND	103	70-130	
Trichlorofluoromethane	6.92	0.50	0.50 ug/L	5.00	ND	138	70-130	QM-05
Trichlorotrifluoroethane	5.94	0.40	0.50 ug/L	5.00	ND	119	70-130	
1,2,3-Trichloropropane	5.18	0.50	0.50 ug/L	5.00	ND	104	70-130	
1,2,4-Trimethylbenzene	5.42	0.50	0.50 ug/L	5.00	ND	108	70-130	
1,3,5-Trimethylbenzene	5.36	0.50	0.50 ug/L	5.00	ND	107	70-130	
Vinyl chloride	6.10	0.50	0.50 ug/L	5.00	ND	122	70-130	
m,p-Xylene	10.5	0.50	0.50 ug/L	10.0	ND	105	70-130	
o-Xylene	5.28	0.50	0.50 ug/L	5.00	ND	106	70-130	
Xylenes (total)	15.7	0.50	0.50 ug/L	15.0	ND	105	70-130	
Methyl tert-butyl ether	4.01	0.50	3.0 ug/L	5.00	ND	80.2	70-130	
Ethyl tert-butyl ether	5.68	0.40	0.50 ug/L	5.00	ND	114	70-130	
Tert-amyl methyl ether	4.20	0.30	0.50 ug/L	5.00	ND	84.0	70-130	
Surrogate: Bromofluorobenzene	26.5		ug/L	25.0		106	70-130	
Surrogate: Dibromofluoromethane	24.6		ug/L	25.0		98.3	70-130	
Surrogate: Toluene-d8	27.1		ug/L	25.0		108	70-130	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

Matrix Spike (AL84158-MS2)	Source: 18L1866-01	Prepared: 12/19/18	Analyzed: 12/20/18	QM-05			
Acetone	28.3	2.0	5.0 ug/L	20.0	ND	141	70-130
Acrylonitrile	6.76	0.40	5.0 ug/L	5.00	ND	135	70-130
Benzene	5.55	0.30	0.30 ug/L	5.00	ND	111	70-130
Bromobenzene	5.34	0.20	0.50 ug/L	5.00	ND	107	70-130
Bromochloromethane	5.22	0.40	0.50 ug/L	5.00	ND	104	70-130
Bromodichloromethane	7.00	0.20	0.50 ug/L	5.00	ND	140	70-130
Bromoform	5.32	0.30	0.50 ug/L	5.00	ND	106	70-130
Bromomethane	6.08	0.40	0.50 ug/L	5.00	ND	122	70-130
n-Butylbenzene	6.06	0.50	0.50 ug/L	5.00	ND	121	70-130
sec-Butylbenzene	6.09	0.50	0.50 ug/L	5.00	ND	122	70-130
tert-Butylbenzene	6.03	0.50	0.50 ug/L	5.00	ND	121	70-130
Carbon disulfide	5.20	0.40	0.50 ug/L	5.00	ND	104	70-130
Carbon tetrachloride	6.21	0.30	0.50 ug/L	5.00	ND	124	70-130
Chlorobenzene	5.40	0.20	0.50 ug/L	5.00	ND	108	70-130
Chloroethane	5.87	0.30	0.50 ug/L	5.00	ND	117	70-130
Chloroform	5.57	0.30	0.50 ug/L	5.00	ND	111	70-130
Chloromethane	4.72	0.40	0.50 ug/L	5.00	ND	94.4	70-130
2-Chlorotoluene	5.72	0.20	0.50 ug/L	5.00	ND	114	70-130
4-Chlorotoluene	5.55	0.20	0.50 ug/L	5.00	ND	111	70-130
Dibromochloromethane	6.26	0.30	0.50 ug/L	5.00	ND	125	70-130
1,2-Dibromo-3-chloropropane	4.93	0.50	1.0 ug/L	5.00	ND	98.6	70-130
1,2-Dibromoethane (EDB)	5.75	0.20	0.50 ug/L	5.00	ND	115	70-130
Dibromomethane	5.01	0.20	0.50 ug/L	5.00	ND	100	70-130
1,2-Dichlorobenzene	5.31	0.20	0.50 ug/L	5.00	ND	106	70-130
1,3-Dichlorobenzene	5.46	0.20	0.50 ug/L	5.00	ND	109	70-130
1,4-Dichlorobenzene	5.15	0.20	0.50 ug/L	5.00	ND	103	70-130
trans-1,4-Dichloro-2-butene	5.25	0.90	5.0 ug/L	5.00	ND	105	70-130
Dichlorodifluoromethane	7.27	0.50	0.50 ug/L	5.00	ND	145	70-130
1,1-Dichloroethane	5.63	0.20	0.50 ug/L	5.00	ND	113	70-130
1,2-Dichloroethane	4.92	0.50	0.50 ug/L	5.00	ND	98.4	70-130
1,1-Dichloroethene	5.43	0.30	0.30 ug/L	5.00	ND	109	70-130
cis-1,2-Dichloroethene	5.36	0.40	0.50 ug/L	5.00	ND	107	70-130
trans-1,2-Dichloroethene	5.68	0.40	0.50 ug/L	5.00	ND	114	70-130
1,2-Dichloropropane	5.25	0.20	0.50 ug/L	5.00	ND	105	70-130

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

Matrix Spike (AL84158-MS2)	Source: 18L1866-01	Prepared: 12/19/18	Analyzed: 12/20/18	QM-05				
1,3-Dichloropropane	5.33	0.50	0.50	ug/L	5.00	ND	107	70-130
2,2-Dichloropropane	6.13	0.30	0.50	ug/L	5.00	ND	123	70-130
1,1-Dichloropropene	5.83	0.50	0.50	ug/L	5.00	ND	117	70-130
cis-1,3-Dichloropropene	4.67	0.30	0.50	ug/L	5.00	ND	93.4	70-130
trans-1,3-Dichloropropene	4.80	0.30	0.50	ug/L	5.00	ND	96.0	70-130
2-Hexanone	6.53	0.50	5.0	ug/L	5.00	ND	131	70-130
Ethylbenzene	5.69	0.50	0.50	ug/L	5.00	ND	114	70-130
Hexachlorobutadiene	5.33	0.40	0.50	ug/L	5.00	ND	107	70-130
Isopropylbenzene	6.15	0.50	0.50	ug/L	5.00	ND	123	70-130
p-Isopropyltoluene	6.12	0.50	0.50	ug/L	5.00	ND	122	70-130
Methyl ethyl ketone	14.0	0.60	1.0	ug/L	10.0	ND	140	70-130
Methyl iodide	5.61	0.40	2.0	ug/L	5.00	ND	112	70-130
Methyl isobutyl ketone	12.9	0.80	1.0	ug/L	10.0	ND	129	70-130
Methylene chloride	3.77	0.40	0.50	ug/L	5.00	ND	75.4	70-130
Naphthalene	5.54	0.50	0.50	ug/L	5.00	ND	111	70-130
n-Propylbenzene	5.98	0.50	0.50	ug/L	5.00	ND	120	70-130
Styrene	5.57	0.20	0.50	ug/L	5.00	ND	111	70-130
1,1,1,2-Tetrachloroethane	5.16	0.40	0.50	ug/L	5.00	ND	103	70-130
1,1,2,2-Tetrachloroethane	5.46	0.20	0.50	ug/L	5.00	ND	109	70-130
Tetrachloroethene	5.23	0.50	0.50	ug/L	5.00	ND	105	70-130
Toluene	5.81	0.30	0.50	ug/L	5.00	ND	116	70-130
1,2,3-Trichlorobenzene	5.35	0.40	0.50	ug/L	5.00	ND	107	70-130
1,2,4-Trichlorobenzene	5.06	0.40	0.50	ug/L	5.00	ND	101	70-130
1,1,1-Trichloroethane	6.20	0.40	0.50	ug/L	5.00	ND	124	70-130
1,1,2-Trichloroethane	5.33	0.20	0.50	ug/L	5.00	ND	107	70-130
Trichloroethene	5.58	0.50	0.50	ug/L	5.00	ND	112	70-130
Trichlorofluoromethane	7.59	0.50	0.50	ug/L	5.00	ND	152	70-130
Trichlorotrifluoroethane	6.52	0.40	0.50	ug/L	5.00	ND	130	70-130
1,2,3-Trichloropropane	5.60	0.50	0.50	ug/L	5.00	ND	112	70-130
1,2,4-Trimethylbenzene	5.95	0.50	0.50	ug/L	5.00	ND	119	70-130
1,3,5-Trimethylbenzene	5.88	0.50	0.50	ug/L	5.00	ND	118	70-130
Vinyl chloride	5.69	0.50	0.50	ug/L	5.00	ND	114	70-130
m,p-Xylene	11.3	0.50	0.50	ug/L	10.0	ND	113	70-130
o-Xylene	5.64	0.50	0.50	ug/L	5.00	ND	113	70-130
Xylenes (total)	17.0	0.50	0.50	ug/L	15.0	ND	113	70-130

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 524.2 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL84158 - VOAs in Water GCMS

Matrix Spike (AL84158-MS2)	Source: 18L1866-01	Prepared: 12/19/18	Analyzed: 12/20/18	QM-05			
Methyl tert-butyl ether	4.21	0.50	3.0 ug/L	5.00	ND	84.2	70-130
Ethyl tert-butyl ether	5.97	0.40	0.50 ug/L	5.00	ND	119	70-130
Tert-amyl methyl ether	4.47	0.30	0.50 ug/L	5.00	ND	89.4	70-130
Surrogate: Bromofluorobenzene	26.5		ug/L	25.0		106	70-130
Surrogate: Dibromofluoromethane	24.5		ug/L	25.0		97.9	70-130
Surrogate: Toluene-d8	26.8		ug/L	25.0		107	70-130



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Blank (AL83983-BLK1)

Prepared & Analyzed: 12/17/18

Acetone	ND	3.0	5.0	ug/L							U
Acetonitrile	ND	50	100	ug/L							U
Allyl chloride	ND	0.40	10	ug/L							U
Acrylonitrile	ND	0.40	5.0	ug/L							U
Benzene	ND	0.30	0.30	ug/L							U
Bromobenzene	ND	0.40	0.50	ug/L							U
Bromochloromethane	ND	0.40	0.50	ug/L							U
Bromodichloromethane	ND	0.40	0.50	ug/L							U
Bromoform	ND	0.30	0.50	ug/L							U
Bromomethane	ND	0.40	0.50	ug/L							U
n-Butylbenzene	ND	0.40	0.50	ug/L							U
sec-Butylbenzene	ND	0.40	0.50	ug/L							U
tert-Butylbenzene	ND	0.30	0.50	ug/L							U
Carbon disulfide	ND	0.40	5.0	ug/L							U
Carbon tetrachloride	ND	0.40	0.50	ug/L							U
Chlorobenzene	ND	0.30	0.50	ug/L							U
Chloroethane	ND	0.40	0.50	ug/L							U
Chloroform	ND	0.40	0.50	ug/L							U
Chloromethane	ND	0.40	0.50	ug/L							U
Chloroprene	ND	0.40	1.0	ug/L							U
2-Chlorotoluene	ND	0.40	0.50	ug/L							U
4-Chlorotoluene	ND	0.30	0.50	ug/L							U
Dibromochloromethane	ND	0.40	0.50	ug/L							U
1,2-Dibromo-3-chloropropane	ND	0.60	2.0	ug/L							U
1,2-Dibromoethane (EDB)	ND	0.40	0.50	ug/L							U
Dibromomethane	ND	0.40	0.50	ug/L							U
1,2-Dichlorobenzene	ND	0.40	0.50	ug/L							U
1,3-Dichlorobenzene	ND	0.40	0.50	ug/L							U
1,4-Dichlorobenzene	ND	0.10	0.50	ug/L							U
trans-1,4-Dichloro-2-butene	ND	0.50	5.0	ug/L							U
Dichlorodifluoromethane	ND	0.40	0.50	ug/L							U
1,1-Dichloroethane	ND	0.10	0.50	ug/L							U
1,2-Dichloroethane	ND	0.40	0.50	ug/L							U
1,1-Dichloroethene	ND	0.30	0.50	ug/L							U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Blank (AL83983-BLK1)

Prepared & Analyzed: 12/17/18

cis-1,2-Dichloroethene	ND	0.40	0.50	ug/L							U
trans-1,2-Dichloroethene	ND	0.40	0.50	ug/L							U
1,2-Dichloropropane	ND	0.40	0.50	ug/L							U
1,3-Dichloropropane	ND	0.40	0.50	ug/L							U
2,2-Dichloropropane	ND	0.20	0.50	ug/L							U
1,1-Dichloropropene	ND	0.40	0.50	ug/L							U
cis-1,3-Dichloropropene	ND	0.40	0.50	ug/L							U
trans-1,3-Dichloropropene	ND	0.40	0.50	ug/L							U
Diethyl ether	ND	0.20	1.0	ug/L							U
Di-isopropyl ether	ND	0.40	0.50	ug/L							U
Ethanol	ND	50	50	ug/L							U
Ethyl methacrylate	ND	0.70	10	ug/L							U
Ethylbenzene	ND	0.40	0.50	ug/L							U
Hexachloroethane	ND	0.40	1.0	ug/L							U
Hexachlorobutadiene	ND	0.50	0.50	ug/L							U
Ethyl tert-butyl ether	ND	0.40	0.50	ug/L							U
2-Hexanone	ND	0.50	5.0	ug/L							U
Isopropylbenzene	ND	0.40	0.50	ug/L							U
Isobutanol	ND	40	100	ug/L							U
p-Isopropyltoluene	ND	0.40	0.50	ug/L							U
Methylene chloride	ND	0.50	0.50	ug/L							U
Methacrylonitrile	ND	0.40	1.0	ug/L							U
Methyl ethyl ketone	ND	0.70	1.0	ug/L							U
Methyl iodide	ND	0.40	2.0	ug/L							U
Methyl methacrylate	ND	0.40	1.0	ug/L							U
Methyl isobutyl ketone	ND	0.60	1.0	ug/L							U
Naphthalene	ND	0.50	0.50	ug/L							U
Propionitrile	ND	20	50	ug/L							U
Methyl tert-butyl ether	ND	0.50	0.50	ug/L							U
n-Propylbenzene	ND	0.40	0.50	ug/L							U
Styrene	ND	0.40	0.50	ug/L							U
Tert-amyl methyl ether	ND	0.40	0.50	ug/L							U
Tert-butyl alcohol	ND	6.0	10	ug/L							U
1,1,1,2-Tetrachloroethane	ND	0.40	0.50	ug/L							U
1,1,2,2-Tetrachloroethane	ND	0.30	0.50	ug/L							U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Blank (AL83983-BLK1)

Prepared & Analyzed: 12/17/18

Tetrachloroethene	ND	0.40	0.50	ug/L							U
Tetrahydrofuran	ND	0.40	5.0	ug/L							U
Toluene	ND	0.30	0.30	ug/L							U
1,2,3-Trichlorobenzene	ND	0.50	0.50	ug/L							U
1,2,4-Trichlorobenzene	ND	0.20	0.50	ug/L							U
1,1,1-Trichloroethane	ND	0.40	0.50	ug/L							U
1,1,2-Trichloroethane	ND	0.40	0.50	ug/L							U
Trichloroethene	ND	0.40	0.50	ug/L							U
Trichlorofluoromethane	ND	0.20	0.50	ug/L							U
1,2,3-Trichloropropane	ND	0.40	0.50	ug/L							U
Trichlorotrifluoroethane	ND	0.50	0.50	ug/L							U
1,2,4-Trimethylbenzene	ND	0.40	0.50	ug/L							U
1,3,5-Trimethylbenzene	ND	0.30	0.50	ug/L							U
Vinyl acetate	ND	0.80	1.0	ug/L							U
Vinyl chloride	ND	0.40	0.50	ug/L							U
m,p-Xylene	ND	0.50	0.50	ug/L							U
o-Xylene	ND	0.40	0.50	ug/L							U
Xylenes (total)	ND	0.50	0.50	ug/L							U
Surrogate: Bromofluorobenzene	27.0			ug/L	25.0		108	70-130			
Surrogate: Dibromofluoromethane	24.8			ug/L	25.0		99.1	70-130			
Surrogate: Toluene-d8	28.2			ug/L	25.0		113	70-130			

LCS (AL83983-BS1)

Prepared & Analyzed: 12/17/18

Acetone	63.3	3.0	5.0	ug/L	80.0		79.1	48-124			
Acetonitrile	1900	50	100	ug/L	2000		94.8	70-130			
Acrylonitrile	20.6	0.40	5.0	ug/L	20.0		103	70-130			
Allyl chloride	24.6	0.40	10	ug/L	20.0		123	70-130			
Benzene	20.8	0.30	0.30	ug/L	20.0		104	82-122			
Bromobenzene	21.2	0.40	0.50	ug/L	20.0		106	83-122			
Bromochloromethane	20.6	0.40	0.50	ug/L	20.0		103	83-124			
Bromodichloromethane	19.3	0.40	0.50	ug/L	20.0		96.5	86-135			
Bromoform	20.4	0.30	0.50	ug/L	20.0		102	76-144			
Bromomethane	18.5	0.40	0.50	ug/L	20.0		92.3	69-145			
n-Butylbenzene	21.2	0.40	0.50	ug/L	20.0		106	79-132			
sec-Butylbenzene	21.0	0.40	0.50	ug/L	20.0		105	86-132			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

LCS (AL83983-BS1)

Prepared & Analyzed: 12/17/18

tert-Butylbenzene	22.3	0.30	0.50	ug/L	20.0		111	82-126			
Carbon disulfide	22.0	0.40	5.0	ug/L	20.0		110	70-130			
Carbon tetrachloride	20.5	0.40	0.50	ug/L	20.0		102	77-134			
Chlorobenzene	20.4	0.30	0.50	ug/L	20.0		102	84-119			
Chloroethane	22.8	0.40	0.50	ug/L	20.0		114	68-133			
Chloroform	20.9	0.40	0.50	ug/L	20.0		105	81-122			
Chloromethane	22.6	0.40	0.50	ug/L	20.0		113	63-129			
Chloroprene	22.7	0.40	1.0	ug/L	20.0		114	70-130			
2-Chlorotoluene	21.2	0.40	0.50	ug/L	20.0		106	79-132			
4-Chlorotoluene	21.2	0.30	0.50	ug/L	20.0		106	80-122			
Dibromochloromethane	19.5	0.40	0.50	ug/L	20.0		97.3	83-135			
1,2-Dibromo-3-chloropropane	22.0	0.60	2.0	ug/L	20.0		110	73-128			
1,2-Dibromoethane (EDB)	20.7	0.40	0.50	ug/L	20.0		103	80-120			
Dibromomethane	20.0	0.40	0.50	ug/L	20.0		99.8	82-124			
1,2-Dichlorobenzene	20.6	0.40	0.50	ug/L	20.0		103	84-121			
1,3-Dichlorobenzene	20.7	0.40	0.50	ug/L	20.0		103	80-120			
1,4-Dichlorobenzene	19.5	0.10	0.50	ug/L	20.0		97.7	84-120			
trans-1,4-Dichloro-2-butene	21.7	0.50	5.0	ug/L	20.0		108	70-130			
Dichlorodifluoromethane	22.2	0.40	0.50	ug/L	20.0		111	52-142			
1,1-Dichloroethane	21.0	0.10	0.50	ug/L	20.0		105	81-126			
1,2-Dichloroethane	19.7	0.40	0.50	ug/L	20.0		98.7	77-117			
1,1-Dichloroethene	21.3	0.30	0.50	ug/L	20.0		106	71-151			
cis-1,2-Dichloroethene	20.0	0.40	0.50	ug/L	20.0		99.9	84-131			
trans-1,2-Dichloroethene	20.3	0.40	0.50	ug/L	20.0		101	79-128			
1,2-Dichloropropane	20.7	0.40	0.50	ug/L	20.0		104	82-125			
1,3-Dichloropropane	21.5	0.40	0.50	ug/L	20.0		107	83-120			
2,2-Dichloropropane	18.3	0.20	0.50	ug/L	20.0		91.5	80-125			
1,1-Dichloropropene	21.8	0.40	0.50	ug/L	20.0		109	85-130			
cis-1,3-Dichloropropene	19.7	0.40	0.50	ug/L	20.0		98.6	83-128			
trans-1,3-Dichloropropene	20.7	0.40	0.50	ug/L	20.0		104	67-129			
Diethyl ether	19.3	0.20	1.0	ug/L	20.0		96.3	70-130			
Di-isopropyl ether	21.2	0.40	0.50	ug/L	20.0		106	83-132			
Ethylbenzene	21.7	0.40	0.50	ug/L	20.0		108	84-124			
Ethyl methacrylate	22.0	0.70	10	ug/L	20.0		110	70-130			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

LCS (AL83983-BS1)

Prepared & Analyzed: 12/17/18

Ethanol	912	50	50	ug/L	980		93.0	50-150			
Hexachloroethane	21.3	0.40	1.0	ug/L	20.0		106	70-130			
Ethyl tert-butyl ether	24.1	0.40	0.50	ug/L	20.0		120	74-127			
Hexachlorobutadiene	19.0	0.50	0.50	ug/L	20.0		95.0	75-135			
2-Hexanone	20.9	0.50	5.0	ug/L	20.0		105	70-130			
Isobutanol	2090	40	100	ug/L	2000		104	70-130			
Isopropylbenzene	22.1	0.40	0.50	ug/L	20.0		110	75-116			
p-Isopropyltoluene	22.3	0.40	0.50	ug/L	20.0		111	78-124			
Methacrylonitrile	19.0	0.40	1.0	ug/L	20.0		94.8	70-130			
Methylene chloride	21.9	0.50	0.50	ug/L	20.0		109	72-132			
Methyl ethyl ketone	39.5	0.70	1.0	ug/L	40.0		98.8	58-157			
Methyl iodide	21.8	0.40	2.0	ug/L	20.0		109	56-167			
Methyl isobutyl ketone	43.3	0.60	1.0	ug/L	40.0		108	70-130			
Methyl methacrylate	21.6	0.40	1.0	ug/L	20.0		108	70-130			
Methyl tert-butyl ether	18.0	0.50	0.50	ug/L	20.0		90.0	84-119			
Naphthalene	20.6	0.50	0.50	ug/L	20.0		103	84-134			
Propionitrile	959	20	50	ug/L	1000		95.9	70-130			
n-Propylbenzene	23.1	0.40	0.50	ug/L	20.0		115	75-127			
Styrene	20.8	0.40	0.50	ug/L	20.0		104	80-125			
Tert-amyl methyl ether	23.1	0.40	0.50	ug/L	20.0		115	74-120			
Tert-butyl alcohol	473	6.0	10	ug/L	400		118	66-147			
1,1,1,2-Tetrachloroethane	20.1	0.40	0.50	ug/L	20.0		100	80-132			
1,1,2,2-Tetrachloroethane	21.2	0.30	0.50	ug/L	20.0		106	84-115			
Tetrachloroethene	20.3	0.40	0.50	ug/L	20.0		102	56-156			
Tetrahydrofuran	20.2	0.40	5.0	ug/L	20.0		101	70-130			
Toluene	21.4	0.30	0.30	ug/L	20.0		107	76-137			
1,2,4-Trichlorobenzene	21.0	0.20	0.50	ug/L	20.0		105	84-126			
1,2,3-Trichlorobenzene	21.4	0.50	0.50	ug/L	20.0		107	85-133			
1,1,1-Trichloroethane	24.7	0.40	0.50	ug/L	20.0		123	70-130			
1,1,2-Trichloroethane	21.4	0.40	0.50	ug/L	20.0		107	83-122			
Trichloroethene	22.4	0.40	0.50	ug/L	20.0		112	84-123			
Trichlorofluoromethane	22.0	0.20	0.50	ug/L	20.0		110	74-130			
1,2,3-Trichloropropane	21.7	0.40	0.50	ug/L	20.0		108	78-122			
Trichlorotrifluoroethane	19.3	0.50	0.50	ug/L	20.0		96.7	82-125			
1,2,4-Trimethylbenzene	22.7	0.40	0.50	ug/L	20.0		113	85-127			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

LCS (AL83983-BS1)

Prepared & Analyzed: 12/17/18

1,3,5-Trimethylbenzene	19.9	0.30	0.50	ug/L	20.0		99.4	80-125			
Vinyl acetate	30.2	0.80	1.0	ug/L	40.0		75.4	60-140			
Vinyl chloride	24.4	0.40	0.50	ug/L	20.0		122	70-130			
m,p-Xylene	43.4	0.50	0.50	ug/L	40.0		109	81-124			
o-Xylene	21.7	0.40	0.50	ug/L	20.0		108	80-126			
Xylenes (total)	65.1	0.50	0.50	ug/L	60.0		108	81-126			
Surrogate: Bromofluorobenzene	28.4			ug/L	25.0		113	70-130			
Surrogate: Dibromofluoromethane	26.1			ug/L	25.0		104	70-130			
Surrogate: Toluene-d8	27.4			ug/L	25.0		110	70-130			

LCS Dup (AL83983-BSD1)

Prepared & Analyzed: 12/17/18

Acetone	65.8	3.0	5.0	ug/L	80.0		82.3	48-124	3.90	25	
Acetonitrile	1860	50	100	ug/L	2000		93.1	70-130	1.74	25	
Acrylonitrile	22.1	0.40	5.0	ug/L	20.0		111	70-130	7.02	25	
Allyl chloride	23.8	0.40	10	ug/L	20.0		119	70-130	3.06	25	
Benzene	20.4	0.30	0.30	ug/L	20.0		102	82-122	1.65	25	
Bromobenzene	21.1	0.40	0.50	ug/L	20.0		106	83-122	0.189	25	
Bromochloromethane	20.6	0.40	0.50	ug/L	20.0		103	83-124	0.0971	25	
Bromodichloromethane	19.8	0.40	0.50	ug/L	20.0		98.8	86-135	2.30	25	
Bromoform	21.1	0.30	0.50	ug/L	20.0		106	76-144	3.52	25	
Bromomethane	18.3	0.40	0.50	ug/L	20.0		91.6	69-145	0.707	25	
n-Butylbenzene	21.5	0.40	0.50	ug/L	20.0		108	79-132	1.83	25	
sec-Butylbenzene	21.0	0.40	0.50	ug/L	20.0		105	86-132	0.0951	25	
tert-Butylbenzene	22.1	0.30	0.50	ug/L	20.0		110	82-126	0.947	25	
Carbon disulfide	22.4	0.40	5.0	ug/L	20.0		112	70-130	2.12	30	
Carbon tetrachloride	21.1	0.40	0.50	ug/L	20.0		105	77-134	2.94	25	
Chlorobenzene	20.2	0.30	0.50	ug/L	20.0		101	84-119	0.836	25	
Chloroethane	22.0	0.40	0.50	ug/L	20.0		110	68-133	3.44	25	
Chloroform	20.6	0.40	0.50	ug/L	20.0		103	81-122	1.49	25	
Chloromethane	22.4	0.40	0.50	ug/L	20.0		112	63-129	0.622	25	
Chloroprene	21.4	0.40	1.0	ug/L	20.0		107	70-130	5.98	25	
2-Chlorotoluene	20.8	0.40	0.50	ug/L	20.0		104	79-132	2.09	25	
4-Chlorotoluene	21.3	0.30	0.50	ug/L	20.0		106	80-122	0.376	25	
Dibromochloromethane	20.5	0.40	0.50	ug/L	20.0		102	83-135	5.11	25	
1,2-Dibromo-3-chloropropane	22.8	0.60	2.0	ug/L	20.0		114	73-128	3.88	25	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

LCS Dup (AL83983-BSD1)

Prepared & Analyzed: 12/17/18

1,2-Dibromoethane (EDB)	20.7	0.40	0.50	ug/L	20.0	104	80-120	0.145	25	
Dibromomethane	20.0	0.40	0.50	ug/L	20.0	99.9	82-124	0.150	25	
1,2-Dichlorobenzene	20.5	0.40	0.50	ug/L	20.0	103	84-121	0.292	25	
1,3-Dichlorobenzene	20.6	0.40	0.50	ug/L	20.0	103	80-120	0.582	25	
1,4-Dichlorobenzene	19.8	0.10	0.50	ug/L	20.0	99.0	84-120	1.37	25	
trans-1,4-Dichloro-2-butene	22.5	0.50	5.0	ug/L	20.0	112	70-130	3.53	25	
Dichlorodifluoromethane	22.6	0.40	0.50	ug/L	20.0	113	52-142	1.92	25	
1,1-Dichloroethane	20.7	0.10	0.50	ug/L	20.0	103	81-126	1.73	25	
1,2-Dichloroethane	19.4	0.40	0.50	ug/L	20.0	97.0	77-117	1.69	25	
1,1-Dichloroethene	20.2	0.30	0.50	ug/L	20.0	101	71-151	5.06	25	
cis-1,2-Dichloroethene	20.0	0.40	0.50	ug/L	20.0	99.8	84-131	0.0501	25	
trans-1,2-Dichloroethene	20.2	0.40	0.50	ug/L	20.0	101	79-128	0.197	25	
1,2-Dichloropropane	20.7	0.40	0.50	ug/L	20.0	103	82-125	0.145	25	
1,3-Dichloropropane	21.0	0.40	0.50	ug/L	20.0	105	83-120	2.07	25	
2,2-Dichloropropane	20.2	0.20	0.50	ug/L	20.0	101	80-125	9.72	25	
1,1-Dichloropropene	21.8	0.40	0.50	ug/L	20.0	109	85-130	0.229	25	
cis-1,3-Dichloropropene	20.4	0.40	0.50	ug/L	20.0	102	83-128	3.24	25	
trans-1,3-Dichloropropene	22.3	0.40	0.50	ug/L	20.0	111	67-129	7.40	25	
Diethyl ether	18.7	0.20	1.0	ug/L	20.0	93.6	70-130	2.79	25	
Di-isopropyl ether	20.7	0.40	0.50	ug/L	20.0	104	83-132	2.15	25	
Ethanol	932	50	50	ug/L	980	95.1	50-150	2.15	25	
Ethylbenzene	21.4	0.40	0.50	ug/L	20.0	107	84-124	1.16	25	
Ethyl methacrylate	22.5	0.70	10	ug/L	20.0	112	70-130	2.29	25	
Ethyl tert-butyl ether	25.3	0.40	0.50	ug/L	20.0	126	74-127	4.82	25	
Hexachloroethane	22.7	0.40	1.0	ug/L	20.0	114	70-130	6.59	25	
Hexachlorobutadiene	19.5	0.50	0.50	ug/L	20.0	97.3	75-135	2.39	25	
2-Hexanone	21.3	0.50	5.0	ug/L	20.0	107	70-130	1.80	30	
Isobutanol	2140	40	100	ug/L	2000	107	70-130	2.69	25	
Isopropylbenzene	21.7	0.40	0.50	ug/L	20.0	109	75-116	1.51	25	
p-Isopropyltoluene	22.5	0.40	0.50	ug/L	20.0	112	78-124	0.849	25	
Methylene chloride	19.2	0.50	0.50	ug/L	20.0	96.2	72-132	12.8	25	
Methacrylonitrile	19.7	0.40	1.0	ug/L	20.0	98.6	70-130	3.93	25	
Methyl ethyl ketone	41.4	0.70	1.0	ug/L	40.0	103	58-157	4.55	25	
Methyl iodide	20.8	0.40	2.0	ug/L	20.0	104	56-167	4.56	30	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

LCS Dup (AL83983-BSD1)

Prepared & Analyzed: 12/17/18

Methyl isobutyl ketone	42.7	0.60	1.0	ug/L	40.0		107	70-130	1.30	25	
Methyl methacrylate	21.8	0.40	1.0	ug/L	20.0		109	70-130	1.11	25	
Propionitrile	1020	20	50	ug/L	1000		102	70-130	6.20	25	
Naphthalene	21.3	0.50	0.50	ug/L	20.0		106	84-134	2.96	25	
Methyl tert-butyl ether	18.6	0.50	0.50	ug/L	20.0		93.2	84-119	3.50	25	
n-Propylbenzene	23.2	0.40	0.50	ug/L	20.0		116	75-127	0.562	25	
Styrene	20.5	0.40	0.50	ug/L	20.0		102	80-125	1.50	25	
Tert-amyl methyl ether	24.1	0.40	0.50	ug/L	20.0		121	74-120	4.32	25	QL-03
Tert-butyl alcohol	472	6.0	10	ug/L	400		118	66-147	0.0360	25	
1,1,1,2-Tetrachloroethane	20.8	0.40	0.50	ug/L	20.0		104	80-132	3.33	25	
1,1,2,2-Tetrachloroethane	21.1	0.30	0.50	ug/L	20.0		106	84-115	0.425	25	
Tetrachloroethene	20.2	0.40	0.50	ug/L	20.0		101	56-156	0.939	25	
Tetrahydrofuran	20.7	0.40	5.0	ug/L	20.0		103	70-130	2.20	25	
Toluene	21.1	0.30	0.30	ug/L	20.0		106	76-137	1.27	25	
1,2,4-Trichlorobenzene	21.7	0.20	0.50	ug/L	20.0		109	84-126	3.56	25	
1,2,3-Trichlorobenzene	22.0	0.50	0.50	ug/L	20.0		110	85-133	2.95	25	
1,1,1-Trichloroethane	25.7	0.40	0.50	ug/L	20.0		128	70-130	3.93	25	
1,1,2-Trichloroethane	21.2	0.40	0.50	ug/L	20.0		106	83-122	0.939	25	
Trichloroethene	22.0	0.40	0.50	ug/L	20.0		110	84-123	2.07	25	
Trichlorofluoromethane	21.0	0.20	0.50	ug/L	20.0		105	74-130	4.61	25	
1,2,3-Trichloropropane	21.5	0.40	0.50	ug/L	20.0		107	78-122	0.972	25	
Trichlorotrifluoroethane	19.3	0.50	0.50	ug/L	20.0		96.3	82-125	0.415	25	
1,2,4-Trimethylbenzene	22.7	0.40	0.50	ug/L	20.0		113	85-127	0.00	25	
1,3,5-Trimethylbenzene	19.7	0.30	0.50	ug/L	20.0		98.7	80-125	0.757	25	
Vinyl acetate	33.5	0.80	1.0	ug/L	40.0		83.7	60-140	10.4	25	
Vinyl chloride	23.6	0.40	0.50	ug/L	20.0		118	70-130	3.16	25	
m,p-Xylene	43.1	0.50	0.50	ug/L	40.0		108	81-124	0.624	25	
o-Xylene	21.3	0.40	0.50	ug/L	20.0		106	80-126	1.68	25	
Xylenes (total)	64.4	0.50	0.50	ug/L	60.0		107	81-126	0.973	25	
Surrogate: Bromofluorobenzene	28.1			ug/L	25.0		112	70-130			
Surrogate: Dibromofluoromethane	26.1			ug/L	25.0		104	70-130			
Surrogate: Toluene-d8	27.0			ug/L	25.0		108	70-130			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Matrix Spike (AL83983-MS1)	Source: 18L1868-01		Prepared: 12/17/18		Analyzed: 12/18/18		QM-05	
Acetone	81.6	3.0	5.0	ug/L	80.0	ND	102	32-164
Acetonitrile	2340	50	100	ug/L	2000	ND	117	70-130
Allyl chloride	22.8	0.40	10	ug/L	20.0	ND	114	70-130
Acrylonitrile	25.2	0.40	5.0	ug/L	20.0	ND	126	70-130
Benzene	21.4	0.30	0.30	ug/L	20.0	ND	107	58-139
Bromobenzene	20.7	0.40	0.50	ug/L	20.0	ND	103	63-143
Bromochloromethane	20.9	0.40	0.50	ug/L	20.0	ND	105	60-141
Bromodichloromethane	18.8	0.40	0.50	ug/L	20.0	ND	94.0	62-140
Bromoform	20.2	0.30	0.50	ug/L	20.0	ND	101	47-165
Bromomethane	17.0	0.40	0.50	ug/L	20.0	ND	85.0	30-163
n-Butylbenzene	17.8	0.40	0.50	ug/L	20.0	ND	88.8	57-147
sec-Butylbenzene	18.4	0.40	0.50	ug/L	20.0	ND	92.2	64-155
tert-Butylbenzene	20.1	0.30	0.50	ug/L	20.0	ND	100	57-150
Carbon disulfide	21.2	0.40	5.0	ug/L	20.0	ND	106	70-130
Carbon tetrachloride	20.8	0.40	0.50	ug/L	20.0	ND	104	65-153
Chlorobenzene	20.3	0.30	0.50	ug/L	20.0	ND	102	58-137
Chloroethane	22.8	0.40	0.50	ug/L	20.0	ND	114	59-141
Chloroform	21.0	0.40	0.50	ug/L	20.0	ND	105	36-151
Chloromethane	22.2	0.40	0.50	ug/L	20.0	ND	111	69-149
Chloroprene	23.8	0.40	1.0	ug/L	20.0	ND	119	70-130
2-Chlorotoluene	20.4	0.40	0.50	ug/L	20.0	ND	102	54-150
4-Chlorotoluene	20.4	0.30	0.50	ug/L	20.0	ND	102	59-140
Dibromochloromethane	19.2	0.40	0.50	ug/L	20.0	ND	96.0	54-157
1,2-Dibromo-3-chloropropane	28.7	0.60	2.0	ug/L	20.0	ND	144	54-137
1,2-Dibromoethane (EDB)	21.3	0.40	0.50	ug/L	20.0	ND	107	40-147
Dibromomethane	20.3	0.40	0.50	ug/L	20.0	ND	102	59-139
1,2-Dichlorobenzene	19.7	0.40	0.50	ug/L	20.0	ND	98.3	39-145
1,3-Dichlorobenzene	18.9	0.40	0.50	ug/L	20.0	ND	94.4	54-137
1,4-Dichlorobenzene	18.6	0.10	0.50	ug/L	20.0	ND	93.2	41-142
trans-1,4-Dichloro-2-butene	26.2	0.50	5.0	ug/L	20.0	ND	131	70-130
Dichlorodifluoromethane	25.2	0.40	0.50	ug/L	20.0	ND	126	39-162
1,1-Dichloroethane	21.8	0.10	0.50	ug/L	20.0	ND	109	39-146
1,2-Dichloroethane	20.2	0.40	0.50	ug/L	20.0	ND	101	58-133
1,1-Dichloroethene	22.9	0.30	0.50	ug/L	20.0	ND	114	70-154

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Matrix Spike (AL83983-MS1)	Source: 18L1868-01	Prepared: 12/17/18	Analyzed: 12/18/18	QM-05			
cis-1,2-Dichloroethene	20.4	0.40	0.50 ug/L	20.0	ND	102	66-141
trans-1,2-Dichloroethene	20.5	0.40	0.50 ug/L	20.0	ND	102	59-151
1,2-Dichloropropane	20.9	0.40	0.50 ug/L	20.0	ND	104	41-142
1,3-Dichloropropane	21.8	0.40	0.50 ug/L	20.0	ND	109	62-139
2,2-Dichloropropane	18.9	0.20	0.50 ug/L	20.0	ND	94.4	40-167
1,1-Dichloropropene	23.8	0.40	0.50 ug/L	20.0	ND	119	58-148
cis-1,3-Dichloropropene	19.2	0.40	0.50 ug/L	20.0	ND	96.2	50-140
trans-1,3-Dichloropropene	20.1	0.40	0.50 ug/L	20.0	ND	101	40-144
Diethyl ether	19.9	0.20	1.0 ug/L	20.0	ND	99.7	70-130
Di-isopropyl ether	21.8	0.40	0.50 ug/L	20.0	ND	109	49-143
Ethanol	1160	50	50 ug/L	980	ND	119	50-150
Ethyl methacrylate	23.7	0.70	10 ug/L	20.0	ND	118	70-130
Ethylbenzene	21.9	0.40	0.50 ug/L	20.0	ND	109	59-147
Ethyl tert-butyl ether	23.8	0.40	0.50 ug/L	20.0	ND	119	44-143
Hexachlorobutadiene	12.2	0.50	0.50 ug/L	20.0	ND	61.2	56-149
Hexachloroethane	19.0	0.40	1.0 ug/L	20.0	ND	95.0	70-130
2-Hexanone	32.7	0.50	5.0 ug/L	20.0	ND	163	70-130
Isobutanol	3140	40	100 ug/L	2000	ND	157	70-130
Isopropylbenzene	21.6	0.40	0.50 ug/L	20.0	ND	108	56-134
p-Isopropyltoluene	19.1	0.40	0.50 ug/L	20.0	ND	95.5	54-148
Methacrylonitrile	22.6	0.40	1.0 ug/L	20.0	ND	113	70-130
Methylene chloride	21.4	0.50	0.50 ug/L	20.0	ND	107	43-143
Methyl ethyl ketone	55.3	0.70	1.0 ug/L	40.0	ND	138	62-126
Methyl iodide	21.2	0.40	2.0 ug/L	20.0	ND	106	70-130
Methyl isobutyl ketone	60.0	0.60	1.0 ug/L	40.0	ND	150	66-127
Methyl methacrylate	25.0	0.40	1.0 ug/L	20.0	ND	125	70-130
Naphthalene	24.0	0.50	0.50 ug/L	20.0	ND	120	52-157
Propionitrile	1190	20	50 ug/L	1000	ND	119	70-130
Methyl tert-butyl ether	17.2	0.50	0.50 ug/L	20.0	ND	86.2	55-144
n-Propylbenzene	21.9	0.40	0.50 ug/L	20.0	ND	109	55-145
Styrene	20.4	0.40	0.50 ug/L	20.0	ND	102	51-157
Tert-amyl methyl ether	21.3	0.40	0.50 ug/L	20.0	ND	107	41-136
Tert-butyl alcohol	832	6.0	10 ug/L	400	ND	208	38-175
1,1,1,2-Tetrachloroethane	19.5	0.40	0.50 ug/L	20.0	ND	97.3	58-146
1,1,2,2-Tetrachloroethane	24.9	0.30	0.50 ug/L	20.0	ND	124	73-127

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Matrix Spike (AL83983-MS1) Source: 18L1868-01 Prepared: 12/17/18 Analyzed: 12/18/18 **QM-05**

Tetrachloroethene	20.8	0.40	0.50	ug/L	20.0	ND	104	49-148			
Tetrahydrofuran	25.8	0.40	5.0	ug/L	20.0	ND	129	70-130			
Toluene	22.0	0.30	0.30	ug/L	20.0	ND	110	59-147			
1,2,4-Trichlorobenzene	17.2	0.20	0.50	ug/L	20.0	ND	86.2	50-150			
1,2,3-Trichlorobenzene	19.5	0.50	0.50	ug/L	20.0	ND	97.6	50-161			
1,1,1-Trichloroethane	25.3	0.40	0.50	ug/L	20.0	ND	126	38-164			
1,1,2-Trichloroethane	21.8	0.40	0.50	ug/L	20.0	ND	109	46-136			
Trichloroethene	21.5	0.40	0.50	ug/L	20.0	ND	108	58-140			
Trichlorofluoromethane	23.1	0.20	0.50	ug/L	20.0	ND	116	56-144			
1,2,3-Trichloropropane	25.0	0.40	0.50	ug/L	20.0	ND	125	61-139			
Trichlorotrifluoroethane	21.9	0.50	0.50	ug/L	20.0	ND	109	59-139			
1,2,4-Trimethylbenzene	21.0	0.40	0.50	ug/L	20.0	ND	105	58-152			
1,3,5-Trimethylbenzene	18.2	0.30	0.50	ug/L	20.0	ND	91.0	58-148			
Vinyl acetate	48.2	0.80	1.0	ug/L	40.0	ND	120	70-130			
Vinyl chloride	24.8	0.40	0.50	ug/L	20.0	ND	124	53-160			
m,p-Xylene	43.6	0.50	0.50	ug/L	40.0	ND	109	53-147			
o-Xylene	21.6	0.40	0.50	ug/L	20.0	ND	108	55-148			
Xylenes (total)	65.2	0.50	0.50	ug/L	60.0	ND	109	49-153			
Surrogate: Bromofluorobenzene	28.1			ug/L	25.0		112	70-130			
Surrogate: Dibromofluoromethane	26.0			ug/L	25.0		104	70-130			
Surrogate: Toluene-d8	27.3			ug/L	25.0		109	70-130			

Matrix Spike (AL83983-MS2) Source: 18L1869-01 Prepared: 12/17/18 Analyzed: 12/18/18

Acetone	69.6	3.0	5.0	ug/L	80.0	ND	87.0	32-164			
Acetonitrile	2310	50	100	ug/L	2000	ND	116	70-130			
Allyl chloride	23.0	0.40	10	ug/L	20.0	ND	115	70-130			
Acrylonitrile	20.0	0.40	5.0	ug/L	20.0	ND	100	70-130			
Benzene	18.6	0.30	0.30	ug/L	20.0	ND	92.8	58-139			
Bromobenzene	17.4	0.40	0.50	ug/L	20.0	ND	87.0	63-143			
Bromochloromethane	17.8	0.40	0.50	ug/L	20.0	ND	88.8	60-141			
Bromodichloromethane	16.6	0.40	0.50	ug/L	20.0	ND	83.2	62-140			
Bromoform	17.5	0.30	0.50	ug/L	20.0	ND	87.6	47-165			
Bromomethane	14.5	0.40	0.50	ug/L	20.0	ND	72.4	30-163			
n-Butylbenzene	19.0	0.40	0.50	ug/L	20.0	ND	95.0	57-147			
sec-Butylbenzene	18.6	0.40	0.50	ug/L	20.0	ND	93.0	64-155			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Matrix Spike (AL83983-MS2)	Source: 18L1869-01			Prepared: 12/17/18 Analyzed: 12/18/18						
tert-Butylbenzene	19.3	0.30	0.50	ug/L	20.0	ND	96.4	57-150		
Carbon disulfide	20.8	0.40	5.0	ug/L	20.0	ND	104	70-130		
Carbon tetrachloride	19.9	0.40	0.50	ug/L	20.0	ND	99.4	65-153		
Chlorobenzene	17.4	0.30	0.50	ug/L	20.0	ND	87.2	58-137		
Chloroethane	21.7	0.40	0.50	ug/L	20.0	ND	109	59-141		
Chloroform	18.6	0.40	0.50	ug/L	20.0	ND	93.1	36-151		
Chloromethane	19.6	0.40	0.50	ug/L	20.0	ND	98.2	69-149		
Chloroprene	20.9	0.40	1.0	ug/L	20.0	ND	104	70-130		
2-Chlorotoluene	18.0	0.40	0.50	ug/L	20.0	ND	89.8	54-150		
4-Chlorotoluene	17.9	0.30	0.50	ug/L	20.0	ND	89.6	59-140		
Dibromochloromethane	16.3	0.40	0.50	ug/L	20.0	ND	81.6	54-157		
1,2-Dibromo-3-chloropropane	19.6	0.60	2.0	ug/L	20.0	ND	98.0	54-137		
1,2-Dibromoethane (EDB)	17.2	0.40	0.50	ug/L	20.0	ND	86.0	40-147		
Dibromomethane	17.5	0.40	0.50	ug/L	20.0	ND	87.6	59-139		
1,2-Dichlorobenzene	17.1	0.40	0.50	ug/L	20.0	ND	85.6	39-145		
1,3-Dichlorobenzene	16.9	0.40	0.50	ug/L	20.0	ND	84.4	54-137		
1,4-Dichlorobenzene	16.4	0.10	0.50	ug/L	20.0	ND	82.0	41-142		
trans-1,4-Dichloro-2-butene	18.3	0.50	5.0	ug/L	20.0	ND	91.3	70-130		
Dichlorodifluoromethane	22.4	0.40	0.50	ug/L	20.0	ND	112	39-162		
1,1-Dichloroethane	19.2	0.10	0.50	ug/L	20.0	ND	95.8	39-146		
1,2-Dichloroethane	16.8	0.40	0.50	ug/L	20.0	ND	84.2	58-133		
1,1-Dichloroethene	19.7	0.30	0.50	ug/L	20.0	ND	98.4	70-154		
cis-1,2-Dichloroethene	17.6	0.40	0.50	ug/L	20.0	ND	87.8	66-141		
trans-1,2-Dichloroethene	18.8	0.40	0.50	ug/L	20.0	ND	94.2	59-151		
1,2-Dichloropropane	17.9	0.40	0.50	ug/L	20.0	ND	89.4	41-142		
1,3-Dichloropropane	17.9	0.40	0.50	ug/L	20.0	ND	89.6	62-139		
2,2-Dichloropropane	19.6	0.20	0.50	ug/L	20.0	ND	97.8	40-167		
1,1-Dichloropropene	21.1	0.40	0.50	ug/L	20.0	ND	105	58-148		
cis-1,3-Dichloropropene	15.7	0.40	0.50	ug/L	20.0	ND	78.6	50-140		
trans-1,3-Dichloropropene	17.2	0.40	0.50	ug/L	20.0	ND	86.0	40-144		
Diethyl ether	16.3	0.20	1.0	ug/L	20.0	ND	81.6	70-130		
Di-isopropyl ether	18.2	0.40	0.50	ug/L	20.0	ND	91.2	49-143		
Ethyl methacrylate	19.0	0.70	10	ug/L	20.0	ND	94.8	70-130		
Ethylbenzene	19.2	0.40	0.50	ug/L	20.0	ND	95.8	59-147		

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AL83983 - VOAs in Water GCMS

Matrix Spike (AL83983-MS2)	Source: 18L1869-01			Prepared: 12/17/18 Analyzed: 12/18/18							
Ethanol	1270	50	50	ug/L	980	ND	129	50-150			
Ethyl tert-butyl ether	20.0	0.40	0.50	ug/L	20.0	ND	99.8	44-143			
Hexachloroethane	18.8	0.40	1.0	ug/L	20.0	ND	94.2	70-130			
Hexachlorobutadiene	15.8	0.50	0.50	ug/L	20.0	ND	79.2	56-149			
2-Hexanone	18.4	0.50	5.0	ug/L	20.0	ND	91.9	70-130			
Isopropylbenzene	19.5	0.40	0.50	ug/L	20.0	ND	97.6	56-134			
Isobutanol	2310	40	100	ug/L	2000	ND	116	70-130			
p-Isopropyltoluene	19.5	0.40	0.50	ug/L	20.0	ND	97.4	54-148			
Methylene chloride	16.5	0.50	0.50	ug/L	20.0	ND	82.7	43-143			
Methacrylonitrile	17.6	0.40	1.0	ug/L	20.0	ND	88.0	70-130			
Methyl ethyl ketone	39.0	0.70	1.0	ug/L	40.0	ND	97.6	62-126			
Methyl iodide	16.7	0.40	2.0	ug/L	20.0	ND	83.6	70-130			
Methyl isobutyl ketone	35.8	0.60	1.0	ug/L	40.0	ND	89.6	66-127			
Methyl methacrylate	18.9	0.40	1.0	ug/L	20.0	ND	94.6	70-130			
Propionitrile	1020	20	50	ug/L	1000	ND	102	70-130			
Methyl tert-butyl ether	14.7	0.50	0.50	ug/L	20.0	ND	73.4	55-144			
Naphthalene	17.9	0.50	0.50	ug/L	20.0	ND	89.4	52-157			
n-Propylbenzene	20.3	0.40	0.50	ug/L	20.0	ND	102	55-145			
Styrene	17.3	0.40	0.50	ug/L	20.0	ND	86.4	51-157			
Tert-amyl methyl ether	18.0	0.40	0.50	ug/L	20.0	ND	89.8	41-136			
Tert-butyl alcohol	552	6.0	10	ug/L	400	ND	138	38-175			
1,1,1,2-Tetrachloroethane	17.0	0.40	0.50	ug/L	20.0	ND	84.8	58-146			
1,1,2,2-Tetrachloroethane	19.2	0.30	0.50	ug/L	20.0	ND	96.1	73-127			
Tetrachloroethene	18.5	0.40	0.50	ug/L	20.0	ND	92.7	49-148			
Tetrahydrofuran	18.1	0.40	5.0	ug/L	20.0	ND	90.4	70-130			
Toluene	19.1	0.30	0.30	ug/L	20.0	ND	95.4	59-147			
1,2,3-Trichlorobenzene	17.9	0.50	0.50	ug/L	20.0	ND	89.7	50-161			
1,2,4-Trichlorobenzene	16.9	0.20	0.50	ug/L	20.0	ND	84.7	50-150			
1,1,1-Trichloroethane	23.6	0.40	0.50	ug/L	20.0	ND	118	38-164			
1,1,2-Trichloroethane	17.9	0.40	0.50	ug/L	20.0	ND	89.6	46-136			
Trichloroethene	19.4	0.40	0.50	ug/L	20.0	ND	96.8	58-140			
Trichlorofluoromethane	21.4	0.20	0.50	ug/L	20.0	ND	107	56-144			
1,2,3-Trichloropropane	18.3	0.40	0.50	ug/L	20.0	ND	91.4	61-139			
Trichlorotrifluoroethane	19.3	0.50	0.50	ug/L	20.0	ND	96.5	59-139			
1,2,4-Trimethylbenzene	19.4	0.40	0.50	ug/L	20.0	ND	97.0	58-152			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Volatile Organic Compounds by EPA Method 8260B - Quality Control

Analyte	Result	MDL	Reporting		Spike Level	Source		%REC		RPD		Notes
			Limit	Units		Result	%REC	Limits	RPD	Limit		

Batch AL83983 - VOAs in Water GCMS

Matrix Spike (AL83983-MS2)

Source: 18L1869-01

Prepared: 12/17/18 Analyzed: 12/18/18

1,3,5-Trimethylbenzene	17.0	0.30	0.50	ug/L	20.0	ND	84.9	58-148			
Vinyl acetate	39.2	0.80	1.0	ug/L	40.0	ND	97.9	70-130			
Vinyl chloride	23.5	0.40	0.50	ug/L	20.0	ND	118	53-160			
m,p-Xylene	37.9	0.50	0.50	ug/L	40.0	ND	94.8	53-147			
o-Xylene	18.6	0.40	0.50	ug/L	20.0	ND	93.0	55-148			
Xylenes (total)	56.5	0.50	0.50	ug/L	60.0	ND	94.2	49-153			
Surrogate: Bromofluorobenzene	27.9			ug/L	25.0		112	70-130			
Surrogate: Dibromofluoromethane	26.3			ug/L	25.0		105	70-130			
Surrogate: Toluene-d8	27.2			ug/L	25.0		109	70-130			



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 262 Rickenbacker Circle, Livermore, CA 94551 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

North Bay: 110 Liberty Street, Petaluma, CA 94952 • Phone: (707) 769-3128 • Fax: (707) 769-8093

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 40ml Amber VOA (NP)
Project Number: Silicone Batch Number 2018090103

Reported:
12/28/18 09:58

Notes and Definitions

- J Detected but below the Reporting Limit; therefore, result is an estimated concentration, detected but not quantified (DNQ).
- QL-03 Although the LCS/LCSD recovery for this analyte is outside of in-house developed control limits, it is within the EPA recommended range of 70-130%.
- QM-05 The spike recovery was outside acceptance limits for the MS and/or MSD due to matrix interference. The LCS and/or LCSD were within acceptance limits showing that the laboratory is in control and the data is acceptable.
- U Analyte included in analysis, but not detected at or above MDL.
- ND Analyte NOT DETECTED at or above the reporting limit
- dry Sample results reported on a dry weight basis
- MDL Method detection limit
- Rec Recovery
- RPD Relative Percent Difference



Laboratory & Corporate: 208 Mason Street, Ukiah, CA 95482
707-468-0401 Fax: 707-468-5267

Service Center & Micro Lab: 262 Rickenbacker Circle, Livermore CA 94551
925-828-6226 Fax: 925-828-6309

Chain of Custody Record

Reports and Invoices will be delivered by email in .pdf format.

Lab No. 18L1885 Page of

<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td colspan="6">Report to: Company: Sample Traps LLC Attn: Quality Control Manager Address: Phone/Fax: Email Address: admin@sampletraps.com</td> <td colspan="6">Invoice to (if different): Company: Attn: Address: Phone/Fax: Email Address:</td> <td colspan="6">Project Info for Report: Project ID: QC- 40ml Amber VOA Vial (NP) Project No: Silicone Batch Number 2018090103 PO/Reference :</td> <td colspan="12" style="text-align: center;">Signature below authorizes work under terms stated on reverse side.</td> </tr> <tr> <td colspan="2">Samplers Signature:</td> <td colspan="2">Print:</td> <td colspan="2">Container:</td> <td colspan="2">Preservative:</td> <td colspan="2">Matrix:</td> <td rowspan="12" style="writing-mode: vertical-rl; transform: rotate(180deg); text-align: center;">Total Number of Containers</td> <td colspan="2">8260 Sample Traps</td> <td colspan="2">524.2 Sample Traps</td> <td colspan="2" rowspan="2" style="writing-mode: vertical-rl; transform: rotate(180deg); text-align: center;">j-flags include chromatograph with report</td> <td rowspan="12" style="writing-mode: vertical-rl; transform: rotate(180deg); text-align: center;">Lab Approval Required For Rush TATs</td> <td colspan="3" style="text-align: center;">Analyses Requested</td> <td colspan="3" style="text-align: center;">TAT</td> <td colspan="3" style="text-align: center;">Sample Notes (lab use only)</td> </tr> <tr> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="3" style="text-align: center;"> 10 days <input type="radio"/> RUSH: <input type="radio"/> 5 days <input type="radio"/> 48 hours <input type="radio"/> Other: <input type="radio"/> ___ days <input type="radio"/> </td> <td colspan="3" style="text-align: center;">Temperature: deg. C Shipment Method: Custody Seals: Y / N</td> </tr> <tr> <td colspan="4">Sample Identification</td> <td colspan="2">Sampled: Date Time</td> <td colspan="2">40ml VOA</td> <td colspan="2">Poly</td> <td colspan="2">Glass bottle</td> <td colspan="2">Glass Jar</td> <td colspan="2">HCL</td> <td colspan="2">Methanol</td> <td colspan="2">Na Bisulfate</td> <td colspan="2">Other</td> <td colspan="2">None</td> <td colspan="2">Water</td> <td colspan="2">Container</td> <td colspan="3" style="text-align: center;">Sample Notes or CDPH Source Numbers:</td> </tr> <tr> <td colspan="4">A8337CUBS - 01</td> <td colspan="2">12/17</td> <td colspan="2">x</td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2">x</td> <td colspan="2">x</td> <td colspan="2">2</td> <td colspan="2">x</td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> </tr> <tr> <td colspan="4">A8337CUBS - 02</td> <td colspan="2">12/17</td> <td colspan="2">x</td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2">x</td> <td colspan="2">x</td> <td colspan="2">2</td> <td colspan="2">x</td> <td colspan="2"></td> <td colspan="2"></td> <td colspan="2"></td> </tr> <tr> <td colspan="30" style="text-align: center;">please use Ukiah reagent water for the analysis</td> </tr> <tr> <td colspan="12">Relinquished by: Per BAL</td> <td colspan="12">Received by: [Signature]</td> <td colspan="3">Date: 12-17-18</td> <td colspan="3">Time: 0730</td> <td colspan="6">CDPH Write On EDT Transmission? <input type="radio"/> Yes <input type="radio"/> No</td> </tr> <tr> <td colspan="12"></td> <td colspan="12"></td> <td colspan="3"></td> <td colspan="3"></td> <td colspan="6">State System Number: _____ If "Y" please enter the Source Number(s) in the column above</td> </tr> <tr> <td colspan="12"></td> <td colspan="12"></td> <td colspan="3"></td> <td colspan="3"></td> <td colspan="6">CA Geotracker EDF Report? <input type="radio"/> Yes <input type="radio"/> No</td> </tr> <tr> <td colspan="12"></td> <td colspan="12"></td> <td colspan="3">Global ID:</td> <td colspan="3">Sampling Company Log Code:</td> <td colspan="6">EDF to (Email Address):</td> </tr> <tr> <td colspan="12"></td> <td colspan="12"></td> <td colspan="3">Travel and Site Time:</td> <td colspan="3">Mileage:</td> <td colspan="6">Misc. Supplies:</td> </tr> </table>												Report to: Company: Sample Traps LLC Attn: Quality Control Manager Address: Phone/Fax: Email Address: admin@sampletraps.com						Invoice to (if different): Company: Attn: Address: Phone/Fax: Email Address:						Project Info for Report: Project ID: QC- 40ml Amber VOA Vial (NP) Project No: Silicone Batch Number 2018090103 PO/Reference :						Signature below authorizes work under terms stated on reverse side.												Samplers Signature:		Print:		Container:		Preservative:		Matrix:		Total Number of Containers	8260 Sample Traps		524.2 Sample Traps		j-flags include chromatograph with report		Lab Approval Required For Rush TATs	Analyses Requested			TAT			Sample Notes (lab use only)																	10 days <input type="radio"/> RUSH: <input type="radio"/> 5 days <input type="radio"/> 48 hours <input type="radio"/> Other: <input type="radio"/> ___ days <input type="radio"/>			Temperature: deg. C Shipment Method: Custody Seals: Y / N			Sample Identification				Sampled: Date Time		40ml VOA		Poly		Glass bottle		Glass Jar		HCL		Methanol		Na Bisulfate		Other		None		Water		Container		Sample Notes or CDPH Source Numbers:			A8337CUBS - 01				12/17		x														x		x		2		x								A8337CUBS - 02				12/17		x														x		x		2		x								please use Ukiah reagent water for the analysis																														Relinquished by: Per BAL												Received by: [Signature]												Date: 12-17-18			Time: 0730			CDPH Write On EDT Transmission? <input type="radio"/> Yes <input type="radio"/> No																																				State System Number: _____ If "Y" please enter the Source Number(s) in the column above																																				CA Geotracker EDF Report? <input type="radio"/> Yes <input type="radio"/> No																														Global ID:			Sampling Company Log Code:			EDF to (Email Address):																														Travel and Site Time:			Mileage:			Misc. Supplies:					
Report to: Company: Sample Traps LLC Attn: Quality Control Manager Address: Phone/Fax: Email Address: admin@sampletraps.com						Invoice to (if different): Company: Attn: Address: Phone/Fax: Email Address:						Project Info for Report: Project ID: QC- 40ml Amber VOA Vial (NP) Project No: Silicone Batch Number 2018090103 PO/Reference :						Signature below authorizes work under terms stated on reverse side.																																																																																																																																																																																																																																																																																																																																																																																											
Samplers Signature:		Print:		Container:		Preservative:		Matrix:		Total Number of Containers	8260 Sample Traps		524.2 Sample Traps		j-flags include chromatograph with report		Lab Approval Required For Rush TATs	Analyses Requested			TAT			Sample Notes (lab use only)																																																																																																																																																																																																																																																																																																																																																																																					
																		10 days <input type="radio"/> RUSH: <input type="radio"/> 5 days <input type="radio"/> 48 hours <input type="radio"/> Other: <input type="radio"/> ___ days <input type="radio"/>			Temperature: deg. C Shipment Method: Custody Seals: Y / N																																																																																																																																																																																																																																																																																																																																																																																								
Sample Identification				Sampled: Date Time		40ml VOA		Poly			Glass bottle		Glass Jar		HCL			Methanol		Na Bisulfate		Other		None		Water		Container		Sample Notes or CDPH Source Numbers:																																																																																																																																																																																																																																																																																																																																																																															
A8337CUBS - 01				12/17		x																x		x		2		x																																																																																																																																																																																																																																																																																																																																																																																	
A8337CUBS - 02				12/17		x																x		x		2		x																																																																																																																																																																																																																																																																																																																																																																																	
please use Ukiah reagent water for the analysis																																																																																																																																																																																																																																																																																																																																																																																																													
Relinquished by: Per BAL												Received by: [Signature]												Date: 12-17-18			Time: 0730			CDPH Write On EDT Transmission? <input type="radio"/> Yes <input type="radio"/> No																																																																																																																																																																																																																																																																																																																																																																															
																														State System Number: _____ If "Y" please enter the Source Number(s) in the column above																																																																																																																																																																																																																																																																																																																																																																															
																														CA Geotracker EDF Report? <input type="radio"/> Yes <input type="radio"/> No																																																																																																																																																																																																																																																																																																																																																																															
																								Global ID:			Sampling Company Log Code:			EDF to (Email Address):																																																																																																																																																																																																																																																																																																																																																																															
																								Travel and Site Time:			Mileage:			Misc. Supplies:																																																																																																																																																																																																																																																																																																																																																																															