



Alpha

Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

ELAP Certificates 1551, 2728, and 2922

31 January 2017

Sample Traps, LLC

Attn: Quality Control Manager

6398 Dougherty Road, Suite 36

Dublin, CA 94568

RE: QC- 32oz. Brown Poly

Work Order: 17A1948

Enclosed are the results of analyses for samples received by the laboratory on 01/19/17 21:20. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Chelsea L. Sandelin

Project Manager



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC
6398 Dougherty Road, Suite 36
Dublin CA, 94568

Project Manager: Quality Control Manager
Project: QC- 32oz. Brown Poly
Project Number: Lot Number J7019CUEN

Reported:
01/31/17 11:14

ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
J7019CUEN	17A1948-01	Water	01/19/17 00:00	01/19/17 21:20



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC
6398 Dougherty Road, Suite 36
Dublin CA, 94568

Project Manager: Quality Control Manager
Project: QC- 32oz. Brown Poly
Project Number: Lot Number J7019CUEN

Reported:
01/31/17 11:14

Miscellaneous Physical/Conventional Chemistry Parameters

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									

J7019CUEN (17A1948-01) Water **Sampled: 01/19/17 00:00** **Received: 01/19/17 21:20**

Cyanide (total)	ND	0.0030	0.0030	mg/L	1	AA74290	01/24/17 14:00	01/24/17 18:39	10-204-00-1X	SMP		U
-----------------	----	--------	--------	------	---	---------	----------------	----------------	--------------	-----	--	---



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC
6398 Dougherty Road, Suite 36
Dublin CA, 94568

Project Manager: Quality Control Manager
Project: QC- 32oz. Brown Poly
Project Number: Lot Number J7019CUEN

Reported:
01/31/17 11:14

Miscellaneous Physical/Conventional Chemistry Parameters - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch AA74290 - General Preparation											
Blank (AA74290-BLK1)					Prepared & Analyzed: 01/24/17						
Cyanide (total)	ND	0.0030	0.0030	mg/L							U
LCS (AA74290-BS1)					Prepared & Analyzed: 01/24/17						
Cyanide (total)	0.197	0.0030	0.0030	mg/L	0.200		98.3	85-115			
Duplicate (AA74290-DUP1)					Source: 17A1029-05 Prepared & Analyzed: 01/24/17						
Cyanide (total)	ND	0.0030	0.0030	mg/L		ND				25	U
Matrix Spike (AA74290-MS1)					Source: 17A1029-05 Prepared & Analyzed: 01/24/17						
Cyanide (total)	0.184	0.0030	0.0030	mg/L	0.200	ND	92.0	85-115			
Matrix Spike Dup (AA74290-MSD1)					Source: 17A1029-05 Prepared & Analyzed: 01/24/17						
Cyanide (total)	0.184	0.0030	0.0030	mg/L	0.200	ND	92.2	85-115	0.195	25	



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC

6398 Dougherty Road, Suite 36
Dublin CA, 94568

Project Manager: Quality Control Manager

Project: QC- 32oz. Brown Poly

Project Number: Lot Number J7019CUEN

Reported:

01/31/17 11:14

Notes and Definitions

U	Analyte included in analysis, but not detected at or above MDL.
ND	Analyte NOT DETECTED at or above the reporting limit
dry	Sample results reported on a dry weight basis
MDL	Method detection limit
Rec	Recovery
RPD	Relative Percent Difference

Chain of Custody Record

Reports and Invoices will be delivered by email in .pdf format.

Lab No. 17A1948 of _____ Page _____

Laboratory & Corporate: 208 Mason Street, Ukiah, CA 95482
707-468-0401 Fax: 707-468-5267

Service Center & Micro Lab: 6388 Dougherty Rd, Ste 35, Dublin, CA 94568
925-828-6226 Fax: 925-828-6309



Invoice to (if different): Company: _____ Attn: _____ Address: _____ Phone/Fax: _____ Email Address: _____ Email: admin@sampletraps.com Samplers Signature: _____		Project Info for Report: Project ID: QC-32oz Brown Poly Project No: Lot number J7019CUEN PO/Reference: _____		Signature _____ Signature below authorizes work under terms stated on reverse side.	
Sample Identification J7019CUEN Print: _____ N/A		Container: 40ml VOA <input checked="" type="checkbox"/> Glass bottle <input type="checkbox"/> Glass Jar <input type="checkbox"/> HCL <input type="checkbox"/> Methanol <input type="checkbox"/> Na Bisulfate <input type="checkbox"/> Other <input type="checkbox"/> None <input checked="" type="checkbox"/> Water <input type="checkbox"/> Container <input type="checkbox"/>		Preservative: Matrix: Cyanide Low Level 3ppb <input checked="" type="checkbox"/> Include MDL -flags <input checked="" type="checkbox"/> Total Number of Containers: 2	
Sampled: Date: _____ Time: _____ X X		Analyses Requested		TAT 10 days <input type="checkbox"/> RUSH: 5 days <input type="checkbox"/> 48 hours <input checked="" type="checkbox"/> Other: _____ days <input type="checkbox"/>	
Sample Notes (lab use only) Temperature: <u>3.3</u> deg. C Shipment Method: _____ Custody Seals: Y / N		Lab Approval Required For Rush TATs		Sample Notes or CDPH Source Numbers:	
Relinquished by: <u>[Signature]</u>		Received by: <u>[Signature]</u>		CDPH Write On EDT Transmission? <input type="radio"/> Yes <input type="radio"/> No	
State System Number: _____ If "y" please enter the Source Number(s) in the column above		CA Geotracker EDF Report? <input type="radio"/> Yes <input type="radio"/> No		Global ID: _____ Sampling Company Log Code: _____ EDF to (Email Address): _____ Travel and Site Time: _____ Mileage: _____ Misc. Supplies: _____	