



*Alpha*

Alpha Analytical Laboratories Inc.

e-mail: [clientservices@alpha-labs.com](mailto:clientservices@alpha-labs.com)

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

ELAP Certificates 1551, 2728, and 2922

09 January 2017

Sample Traps, LLC

Attn: Quality Control Manager

6398 Dougherty Road, Suite 36

Dublin, CA 94568

RE: QC- 2.5g Cubitainer

Work Order: 16L2207

Enclosed are the results of analyses for samples received by the laboratory on 12/19/16 13:40. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Jeanette L. Poplin For Chelsea L. Sandelin

Project Manager



Alpha Analytical Laboratories Inc.

e-mail: [clientservices@alpha-labs.com](mailto:clientservices@alpha-labs.com)

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

#### ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
N6337GUCS	16L2207-01	Water	12/19/16 00:00	12/19/16 13:40



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Metals (Drinking Water) by EPA 200 Series Methods

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
<b>N6337GUCS (16L2207-01) Water    Sampled: 12/19/16 00:00    Received: 12/19/16 13:40</b>												
Calcium	ND	0.020	1.0		mg/L	1	AL64319	01/03/17 10:30	01/04/17 13:37	EPA 200.7	MAM	U
Magnesium	ND	0.0080	1.0		mg/L	1	AL64319	01/03/17 10:30	01/04/17 13:37	EPA 200.7	MAM	U
Potassium	ND	0.040	1.0		mg/L	1	AL64319	01/03/17 10:30	01/04/17 13:37	EPA 200.7	MAM	U
Sodium	ND	0.020	1.0		mg/L	1	AL64319	01/03/17 10:30	01/04/17 13:37	EPA 200.7	MAM	U



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Metals by EPA 200 Series Methods

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
<b>N6337GUCS (16L2207-01) Water    Sampled: 12/19/16 00:00    Received: 12/19/16 13:40</b>												
Mercury	ND	0.010	0.020		ug/L	1	AL64078	12/27/16 07:30	12/28/16 12:22	EPA 245.1	JJM	U
Tin	ND	0.020	0.050		mg/L	1	AL64319	01/03/17 10:30	01/04/17 13:37	EPA 200.7	MAM	U



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Metals by EPA Method 200.8 ICP/MS

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
<b>N6337GUCS (16L2207-01) Water    Sampled: 12/19/16 00:00    Received: 12/19/16 13:40</b>												
Aluminum	ND	2.0	10		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Antimony	ND	0.020	0.50		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Arsenic	ND	0.20	0.50		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Barium	ND	0.050	0.50		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Beryllium	ND	0.020	0.10		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
<b>Boron</b>	<b>26</b>	<b>6.0</b>	<b>50</b>		<b>ug/L</b>	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	J
Cadmium	ND	0.050	0.10		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Chromium	ND	0.080	0.50		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Cobalt	ND	0.010	0.10		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Copper	ND	0.050	0.50		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
<b>Iron</b>	<b>14</b>	<b>2.0</b>	<b>50</b>		<b>ug/L</b>	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	J
Lead	ND	0.020	0.25		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Manganese	ND	0.050	5.0		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Molybdenum	ND	0.020	0.25		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Nickel	ND	0.060	0.50		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Selenium	ND	0.20	2.0		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Silver	ND	0.050	0.10		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Thallium	ND	0.050	0.10		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
Vanadium	ND	0.50	1.0		ug/L	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	U
<b>Zinc</b>	<b>0.72</b>	<b>0.50</b>	<b>5.0</b>		<b>ug/L</b>	1	AL64326	01/03/17 10:42	01/06/17 07:17	EPA 200.8	MMY	J



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Anions by EPA Method 300.0

Analyte	Result	MDL	Reporting		Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	Notes
			Limit									
<b>N6337GUCS (16L2207-01) Water    Sampled: 12/19/16 00:00    Received: 12/19/16 13:40</b>												
Fluoride	ND	0.070	0.10		mg/L	1	AL63908	12/21/16 22:45	12/21/16 22:45	EPA 300.0	SMP	U
Nitrate as NO3	ND	0.20	1.0		mg/L	1	AL63908	12/21/16 22:45	12/21/16 22:45	EPA 300.0	SMP	T-2, U
Nitrite as NO2	ND	0.070	1.0		mg/L	1	AL63908	12/21/16 22:45	12/21/16 22:45	EPA 300.0	SMP	T-2, U
Sulfate as SO4	ND	0.20	0.50		mg/L	1	AL63908	12/21/16 22:45	12/21/16 22:45	EPA 300.0	SMP	U



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Metals (Drinking Water) by EPA 200 Series Methods - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

#### Batch AL64319 - Metals Digest

##### Blank (AL64319-BLK1)

Prepared: 01/03/17 Analyzed: 01/04/17

Calcium	0.0253	0.020	1.0	mg/L							J
Magnesium	ND	0.0080	1.0	mg/L							U
Potassium	ND	0.040	1.0	mg/L							U
Sodium	0.0298	0.020	1.0	mg/L							J

##### LCS (AL64319-BS1)

Prepared: 01/03/17 Analyzed: 01/04/17

Calcium	7.04	0.020	1.0	mg/L	7.20		97.8	85-115			
Magnesium	7.14	0.0080	1.0	mg/L	7.40		96.5	85-115			
Potassium	7.04	0.040	1.0	mg/L	7.60		92.7	85-115			
Sodium	7.76	0.020	1.0	mg/L	7.40		105	85-115			

##### Duplicate (AL64319-DUP1)

Source: 16L1930-01

Prepared: 01/03/17 Analyzed: 01/04/17

Calcium	7.14	0.020	1.0	mg/L		7.16			0.171	20	
Magnesium	3.70	0.0080	1.0	mg/L		3.74			0.889	20	
Potassium	1.01	0.040	1.0	mg/L		1.00			0.502	20	
Sodium	19.5	0.020	1.0	mg/L		19.6			0.499	20	

##### Matrix Spike (AL64319-MS1)

Source: 16L1930-01

Prepared: 01/03/17 Analyzed: 01/04/17

Calcium	14.5	0.020	1.0	mg/L	7.20	7.16	101	70-130			
Magnesium	11.9	0.0080	1.0	mg/L	7.40	3.74	110	70-130			
Potassium	8.06	0.040	1.0	mg/L	7.60	1.00	92.8	70-130			
Sodium	26.5	0.020	1.0	mg/L	7.40	19.6	93.7	70-130			

##### Matrix Spike (AL64319-MS2)

Source: 16L2020-01

Prepared: 01/03/17 Analyzed: 01/04/17

Calcium	87.4	0.020	1.0	mg/L	7.20	82.5	68.2	70-130			QM-4X
Magnesium	71.5	0.0080	1.0	mg/L	7.40	65.2	85.2	70-130			
Potassium	8.67	0.040	1.0	mg/L	7.60	1.38	95.9	70-130			
Sodium	51.0	0.020	1.0	mg/L	7.40	44.8	84.2	70-130			



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Metals (Drinking Water) by EPA 200 Series Methods - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	--------------------	-------	----------------	------------------	------	----------------	-----	--------------	-------

#### Batch AL64319 - Metals Digest

##### Matrix Spike Dup (AL64319-MSD1)

Source: 16L1930-01

Prepared: 01/03/17 Analyzed: 01/04/17

Calcium	14.3	0.020	1.0	mg/L	7.20	7.16	99.8	70-130	0.793	20	
Magnesium	11.7	0.0080	1.0	mg/L	7.40	3.74	107	70-130	1.95	20	
Potassium	8.33	0.040	1.0	mg/L	7.60	1.00	96.4	70-130	3.33	20	
Sodium	26.8	0.020	1.0	mg/L	7.40	19.6	97.2	70-130	0.980	20	





Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

**Metals by EPA 200 Series Methods - Quality Control**

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

**Batch AL64078 - EPA 245.1 Hg Water**

<b>Blank (AL64078-BLK1)</b>			Prepared: 12/27/16 Analyzed: 12/28/16								
Mercury	ND	0.010	0.020	ug/L							U
<b>LCS (AL64078-BS1)</b>			Prepared: 12/27/16 Analyzed: 12/28/16								
Mercury	0.265	0.010	0.020	ug/L	0.250		106	85-115			
<b>Duplicate (AL64078-DUP1)</b>			<b>Source: 16L2244-01</b>			Prepared: 12/27/16 Analyzed: 12/28/16					
Mercury	ND	0.010	0.020	ug/L		ND			20		U
<b>Matrix Spike (AL64078-MS1)</b>			<b>Source: 16L2244-01</b>			Prepared: 12/27/16 Analyzed: 12/28/16					
Mercury	0.263	0.010	0.020	ug/L	0.250	ND	105	70-130			
<b>Matrix Spike (AL64078-MS2)</b>			<b>Source: 16L2093-04</b>			Prepared: 12/27/16 Analyzed: 12/28/16					
Mercury	0.267	0.010	0.020	ug/L	0.250	ND	107	70-130			
<b>Matrix Spike Dup (AL64078-MSD1)</b>			<b>Source: 16L2244-01</b>			Prepared: 12/27/16 Analyzed: 12/28/16					
Mercury	0.264	0.010	0.020	ug/L	0.250	ND	106	70-130	0.380	20	

**Batch AL64319 - Metals Digest**

<b>Blank (AL64319-BLK1)</b>			Prepared: 01/03/17 Analyzed: 01/04/17								
Tin	ND	0.020	0.050	mg/L							U
<b>LCS (AL64319-BS1)</b>			Prepared: 01/03/17 Analyzed: 01/04/17								
Tin	0.210	0.020	0.050	mg/L	0.200		105	85-115			
<b>Duplicate (AL64319-DUP1)</b>			<b>Source: 16L1930-01</b>			Prepared: 01/03/17 Analyzed: 01/04/17					
Tin	ND	0.020	0.050	mg/L		ND			20		U



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Metals by EPA 200 Series Methods - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

#### Batch AL64319 - Metals Digest

<b>Matrix Spike (AL64319-MS1)</b>		<b>Source: 16L1930-01</b>			<b>Prepared: 01/03/17 Analyzed: 01/04/17</b>					
Tin	0.193	0.020	0.050	mg/L	0.200	ND	96.7	70-130		
<b>Matrix Spike (AL64319-MS2)</b>		<b>Source: 16L2020-01</b>			<b>Prepared: 01/03/17 Analyzed: 01/04/17</b>					
Tin	0.189	0.020	0.050	mg/L	0.200	ND	94.4	70-130		
<b>Matrix Spike Dup (AL64319-MSD1)</b>		<b>Source: 16L1930-01</b>			<b>Prepared: 01/03/17 Analyzed: 01/04/17</b>					
Tin	0.193	0.020	0.050	mg/L	0.200	ND	96.3	70-130	0.482	20



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Metals by EPA Method 200.8 ICP/MS - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

#### Batch AL64326 - EPA 200.8

##### Blank (AL64326-BLK1)

Prepared: 01/03/17 Analyzed: 01/06/17

Aluminum	ND	2.0	10	ug/L							U
Antimony	ND	0.020	0.50	ug/L							U
Arsenic	ND	0.20	0.50	ug/L							U
Barium	ND	0.050	0.50	ug/L							U
Beryllium	ND	0.020	0.10	ug/L							U
Boron	ND	6.0	50	ug/L							U
Cadmium	ND	0.050	0.10	ug/L							U
Chromium	ND	0.080	0.50	ug/L							U
Cobalt	ND	0.010	0.10	ug/L							U
Copper	0.267	0.050	0.50	ug/L							J
Iron	9.52	2.0	50	ug/L							J
Lead	ND	0.020	0.25	ug/L							U
Manganese	ND	0.050	5.0	ug/L							U
Molybdenum	ND	0.020	0.25	ug/L							U
Nickel	ND	0.060	0.50	ug/L							U
Selenium	ND	0.20	2.0	ug/L							U
Silver	ND	0.050	0.10	ug/L							U
Thallium	ND	0.050	0.10	ug/L							U
Vanadium	ND	0.50	1.0	ug/L							U
Zinc	0.731	0.50	5.0	ug/L							J

##### LCS (AL64326-BS1)

Prepared: 01/03/17 Analyzed: 01/07/17

Aluminum	550	2.0	10	ug/L	520		106	85-115			
Antimony	20.5	0.020	0.50	ug/L	20.0		102	85-115			
Arsenic	22.2	0.20	0.50	ug/L	20.0		111	85-115			
Barium	21.4	0.050	0.50	ug/L	20.0		107	85-115			
Beryllium	21.1	0.020	0.10	ug/L	20.0		105	85-115			
Boron	98.8	6.0	50	ug/L	100		98.8	85-115			
Cadmium	20.8	0.050	0.10	ug/L	20.0		104	85-115			
Chromium	18.3	0.080	0.50	ug/L	20.0		91.4	85-115			
Cobalt	19.2	0.010	0.10	ug/L	20.0		96.2	85-115			
Copper	20.6	0.050	0.50	ug/L	20.0		103	85-115			
Iron	485	2.0	50	ug/L	520		93.3	85-115			
Lead	19.8	0.020	0.25	ug/L	20.0		99.2	85-115			
Manganese	18.0	0.050	5.0	ug/L	20.0		90.0	85-115			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

**Metals by EPA Method 200.8 ICP/MS - Quality Control**

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

**Batch AL64326 - EPA 200.8**

**LCS (AL64326-BS1)**

Prepared: 01/03/17 Analyzed: 01/06/17

Molybdenum	22.8	0.020	0.25	ug/L	20.0	114	85-115				
Nickel	19.5	0.060	0.50	ug/L	20.0	97.6	85-115				
Selenium	20.0	0.20	2.0	ug/L	20.0	100	85-115				
Silver	21.8	0.050	0.10	ug/L	20.0	109	85-115				
Thallium	20.0	0.050	0.10	ug/L	20.0	100	85-115				
Vanadium	18.6	0.50	1.0	ug/L	20.0	93.1	85-115				
Zinc	110	0.50	5.0	ug/L	100	110	85-115				

**Duplicate (AL64326-DUP1)**

Source: 16L1902-02

Prepared: 01/03/17 Analyzed: 01/06/17

Aluminum	29.0	8.0	40	ug/L	25.8			11.7	20		J
Antimony	0.305	0.080	2.0	ug/L	0.166			59.0	20		J
Arsenic	1.75	0.80	2.0	ug/L	1.92			9.15	20		J
Barium	2.69	0.20	2.0	ug/L	2.94			8.63	20		
Beryllium	ND	0.080	0.40	ug/L	ND				20		R-01, U
Boron	54.2	24	200	ug/L	57.1			5.07	20		J
Cadmium	ND	0.20	0.40	ug/L	ND				20		R-01, U
Chromium	ND	0.32	2.0	ug/L	ND				20		R-01, U
Cobalt	0.710	0.040	0.40	ug/L	0.749			5.33	20		
Copper	11.5	0.20	2.0	ug/L	13.4			15.0	20		
Iron	2540	8.0	200	ug/L	2630			3.40	20		
Lead	3.41	0.080	1.0	ug/L	3.59			4.98	20		
Manganese	468	0.20	20	ug/L	485			3.52	20		
Molybdenum	0.464	0.080	1.0	ug/L	0.458			1.21	20		J
Nickel	2.09	0.24	2.0	ug/L	2.18			4.25	20		
Selenium	ND	0.80	8.0	ug/L	ND				20		R-01, U
Silver	ND	0.20	0.40	ug/L	ND				20		R-01, U
Thallium	ND	0.20	0.40	ug/L	ND				20		R-01, U
Vanadium	ND	2.0	4.0	ug/L	ND				20		R-01, U
Zinc	27.4	2.0	20	ug/L	29.8			8.38	20		

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Anions by EPA Method 300.0 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

#### Batch AL63908 - General Preparation

##### Blank (AL63908-BLK1)

Prepared & Analyzed: 12/21/16

Nitrate as NO3	ND	0.20	1.0	mg/L							U
Fluoride	ND	0.070	0.10	mg/L							U
Nitrite as NO2	ND	0.070	1.0	mg/L							U
Sulfate as SO4	ND	0.20	0.50	mg/L							U

##### LCS (AL63908-BS1)

Prepared & Analyzed: 12/21/16

Sulfate as SO4	22.9	0.20	0.50	mg/L	22.2		103	90-110			
Nitrite as NO2	19.2	0.070	1.0	mg/L	18.2		106	90-110			
Nitrate as NO3	26	0.20	1.0	mg/L	24.7		103	90-110			
Fluoride	5.55	0.070	0.10	mg/L	5.56		99.9	90-110			

##### Duplicate (AL63908-DUP1)

Source: 16L2059-01

Prepared & Analyzed: 12/21/16

Sulfate as SO4	37.4	0.20	0.50	mg/L		37.2			0.343	20	
Nitrate as NO3	0.41	0.20	1.0	mg/L		0.41			0.244	20	J
Nitrite as NO2	ND	0.070	1.0	mg/L		ND				20	U
Fluoride	0.0840	0.070	0.10	mg/L		0.0830			1.20	20	J

##### Matrix Spike (AL63908-MS1)

Source: 16L2059-01

Prepared & Analyzed: 12/21/16

Fluoride	5.19	0.35	0.50	mg/L	5.56	ND	93.5	80-120			
Nitrite as NO2	20.1	0.35	5.0	mg/L	18.2	ND	110	80-120			
Sulfate as SO4	60.4	1.0	2.5	mg/L	22.2	37.2	104	80-120			
Nitrate as NO3	24	1.0	5.0	mg/L	24.7	ND	98.5	80-120			

##### Matrix Spike (AL63908-MS2)

Source: 16L2084-01

Prepared & Analyzed: 12/21/16

Nitrite as NO2	19.8	0.070	1.0	mg/L	18.2	ND	109	80-120			
Sulfate as SO4	39.2	0.20	0.50	mg/L	22.2	16.2	103	80-120			
Fluoride	5.90	0.070	0.10	mg/L	5.56	0.179	103	80-120			
Nitrate as NO3	27	0.20	1.0	mg/L	24.7	ND	107	80-120			



Alpha Analytical Laboratories Inc.

e-mail: clientservices@alpha-labs.com

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Anions by EPA Method 300.0 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	--------------------	-------	----------------	------------------	------	----------------	-----	--------------	-------

#### Batch AL63908 - General Preparation

##### Matrix Spike Dup (AL63908-MSD1)

Source: 16L2059-01

Prepared & Analyzed: 12/21/16

Nitrite as NO <sub>2</sub>	20.3	0.35	5.0	mg/L	18.2	ND	112	80-120	1.18	20	
Nitrate as NO <sub>3</sub>	24	1.0	5.0	mg/L	24.7	ND	96.7	80-120	1.80	20	
Sulfate as SO <sub>4</sub>	60.2	1.0	2.5	mg/L	22.2	37.2	104	80-120	0.221	20	
Fluoride	5.15	0.35	0.50	mg/L	5.56	ND	92.7	80-120	0.859	20	



Alpha Analytical Laboratories Inc.

e-mail: [clientservices@alpha-labs.com](mailto:clientservices@alpha-labs.com)

Corporate: 208 Mason St., Ukiah, CA 95482 • Phone: (707) 468-0401 • Fax: (707) 468-5267

Bay Area: 6398 Dougherty Rd., Suite 35, Dublin, CA 94568 • Phone: (925) 828-6226 • Fax: (925) 828-6309

Central Valley: 9090 Union Park Way, Suite 113, Elk Grove, CA 95624 • Phone: (916) 686-5190 • Fax: (916) 686-5192

Sample Traps, LLC  
6398 Dougherty Road, Suite 36  
Dublin CA, 94568

Project Manager: Quality Control Manager  
Project: QC- 2.5g Cubitainer  
Project Number: Lot Number N6337GUCS

Reported:  
01/09/17 12:47

### Notes and Definitions

- J Detected but below the Reporting Limit; therefore, result is an estimated concentration, detected but not quantified (DNQ).
- QM-4X The spike recovery was outside of QC acceptance limits for the MS and/or MSD due to analyte concentration at 4 times or greater the spike concentration. The QC batch was accepted based on LCS and/or LCSD recoveries within the acceptance limits.
- R-01 The Reporting Limit for this analyte has been raised to account for matrix interference.
- T-2 Sample analyzed outside of recommended holding time per client.
- U Analyte included in analysis, but not detected at or above MDL.
- ND Analyte NOT DETECTED at or above the reporting limit
- dry Sample results reported on a dry weight basis
- MDL Method detection limit
- Rec Recovery
- RPD Relative Percent Difference

