



alpha

Alpha Analytical Laboratories, Inc.

email: clientservices@alpha-labs.com

Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

27 October 2021

Sample Traps, LLC

Attn: Quality Control Manager

262 Rickenbacker Circle

Livermore, CA 94551

RE: QC- 32oz Glass Boston Round Amber

Work Order: 21J0119

Enclosed are the results of analyses for samples received by the laboratory on 09/23/21 23:20. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Chelsea L. Sandelin

Project Manager



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 32oz Glass Boston Round Amber
Project Number: Lot Number G1271CUFS

Reported:
10/27/21 11:37

Bay Area: 262 Rickenbacker Circle | Livermore, CA 94551 | T: 925-828-6226 | F: 925-828-6309 | ELAP# 2728
Central Valley: 9090 Union Park Way Suite 113 | Elk Grove, CA 95624 | T: 916-686-5190 | F: 916-686-5192 | ELAP# 2922
North Bay: 110 Liberty Street | Petaluma, CA 94952 | T: 707-769-3128 | F: 707-769-8093 | ELAP# 2303
San Diego: 2722 Loker Avenue West Suite A | Carlsbad, CA 92010 | T: 760-930-2555 | F: 760-930-2510 | ELAP# 3055

ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
G1271CUFS - 01	21J0119-01	Water	10/04/21 00:00	09/23/21 23:20
G1271CUFS - 02	21J0119-02	Water	10/04/21 00:00	09/23/21 23:20
G1271CUFS - 03	21J0119-03	Water	10/04/21 00:00	09/23/21 23:20
G1271CUFS - 04	21J0119-04	Water	10/04/21 00:00	09/23/21 23:20
G1271CUFS - 05	21J0119-05	Water	10/04/21 00:00	09/23/21 23:20



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Organochlorine Pesticides and PCBs by EPA Method 608.3

Analyte	Result	MDL	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	ELAP#	Notes
G1271CUFS - 01 (21J0119-01) Water Sampled: 10/04/21 00:00 Received: 09/23/21 23:20												
Aldrin	ND	0.0020	0.0050	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
alpha-BHC	ND	0.0040	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
beta-BHC	ND	0.0040	0.0050	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
delta-BHC	ND	0.0020	0.0050	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
gamma-BHC (Lindane)	ND	0.0040	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Chlordane (tech)	ND	0.010	0.050	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
4,4'-DDD	ND	0.0050	0.020	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
4,4'-DDE	ND	0.0040	0.020	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
4,4'-DDT	ND	0.0030	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Dieldrin	ND	0.0050	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Endosulfan I	ND	0.0040	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Endosulfan II	ND	0.0050	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Endosulfan sulfate	ND	0.0030	0.050	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Endrin	ND	0.0080	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Endrin aldehyde	ND	0.0040	0.040	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Heptachlor	ND	0.0050	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Heptachlor epoxide	ND	0.0090	0.010	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Methoxychlor	ND	0.010	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Toxaphene	ND	0.20	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
PCB-1016	ND	0.080	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
PCB-1221	ND	0.20	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
PCB-1232	ND	0.10	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
PCB-1242	ND	0.040	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
PCB-1248	ND	0.060	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
PCB-1254	ND	0.040	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
PCB-1260	ND	0.040	0.50	ug/L	1	AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	U
Surrogate: Dibutylchlorodate		87.4 %	10-147			AJ13545	10/11/21 07:00	10/14/21 21:52	EPA 608.3	MCB	1551	



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC
 262 Rickenbacker Circle
 Livermore CA, 94551

Project Manager: Quality Control Manager
 Project: QC- 32oz Glass Boston Round Amber
 Project Number: Lot Number G1271CUFS

Reported:
 10/27/21 11:37

Chlorinated Herbicides by EPA Method 8151A

Analyte	Result	MDL	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	ELAP#	Notes
G1271CUFS - 03 (21J0119-03) Water Sampled: 10/04/21 00:00 Received: 09/23/21 23:20												
2,4-D	ND	1.0	1.0	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
2,4-DB	ND	1.0	4.0	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
2,4,5-T	ND	0.20	0.50	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
2,4,5-TP (Silvex)	ND	0.20	0.50	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
Dalapon	ND	2.0	6.0	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
Dicamba	ND	0.10	0.40	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
Dinoseb	ND	0.20	1.0	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
Dichlorprop	ND	0.40	1.0	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
Pentachlorophenol	ND	0.20	0.20	ug/L	1	AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	U
Surrogate: DCAA		100 %	62-124			AJ13382	10/06/21 06:58	10/08/21 19:14	EPA 8151A	MCB	1551	



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Organophosphorus Compounds by EPA Method 8141A

Analyte	Result	MDL	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	ELAP #	Notes
G1271CUFS - 02 (21J0119-02) Water Sampled: 10/04/21 00:00 Received: 09/23/21 23:20 T-06												
Azinphos ethyl	ND	1.0	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Azinphos methyl	ND	0.40	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Bolstar	ND	0.30	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Chlorpyrifos	ND	0.70	1.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Coumaphos	ND	0.80	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Demeton-o	ND	0.20	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Demeton-s	ND	0.30	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Diazinon	ND	0.30	0.50	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Dichlorvos	ND	0.40	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Dimethoate	ND	1.0	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Disulfoton	ND	0.30	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
EPN	ND	0.50	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Ethion	ND	0.80	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Ethoprop (Ethoprophos)	ND	0.40	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Fensulfothion	ND	0.80	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Fenthion	ND	0.50	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Malathion	ND	0.80	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Mevinphos	ND	0.80	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Parathion	ND	0.90	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Parathion-methyl	ND	0.40	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Phorate	ND	0.40	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Ronnel	ND	0.30	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Stirofos	ND	0.30	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Thionazin	ND	1.0	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Tokuthion (Prothiofos)	ND	0.40	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Trichloronate	ND	0.40	2.0	ug/L	1	AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	U
Surrogate: Tributyl phosphate		84.8 %		37-115		AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	
Surrogate: Triphenyl phosphate		97.4 %		44-145		AJ14235	10/20/21 07:00	10/26/21 03:41	EPA 8141A	MCB	1551	



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C

Analyte	Result	MDL	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	ELAP#	Notes
G1271CUFS - 04 (21J0119-04) Water Sampled: 10/04/21 00:00 Received: 09/23/21 23:20												
Acenaphthene	ND	0.60	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Acenaphthylene	ND	0.50	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Anthracene	ND	0.40	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Benzo (a) anthracene	ND	0.40	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Benzo (b) pyrene	ND	0.50	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Benzo (b) fluoranthene	ND	0.60	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Benzo (g,h,i) perylene	ND	0.90	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Benzo (k) fluoranthene	ND	0.50	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Bis(2-chloroethoxy)methane	ND	0.90	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Bis(2-chloroethyl)ether	ND	0.90	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Bis(2-chloroisopropyl)ether	ND	1.0	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Bis(2-ethylhexyl)phthalate	ND	3.0	5.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
4-Bromophenyl phenyl ether	ND	1.0	4.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Butyl benzyl phthalate	ND	0.60	5.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
4-Chloro-3-methylphenol	ND	1.0	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
4-Chloroaniline	ND	1.0	14	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2-Chloronaphthalene	ND	0.60	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2-Chlorophenol	ND	0.70	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
4-Chlorophenyl phenyl ether	ND	0.90	4.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Chrysene	ND	0.80	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Di-n-butyl phthalate	4.1	0.90	10	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	J
Di-n-octyl phthalate	ND	0.60	10	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Dibenz (a,h) anthracene	ND	0.80	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Dibenzofuran	ND	0.90	4.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
1,2-Dichlorobenzene	ND	0.60	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
1,3-Dichlorobenzene	ND	0.60	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
1,4-Dichlorobenzene	ND	0.60	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2,4-Dichlorophenol	ND	0.70	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Diethyl phthalate	ND	0.30	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Dimethyl phthalate	ND	0.80	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2,4-Dimethylphenol	ND	1.0	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
4,6-Dinitro-2-methylphenol	ND	3.0	4.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2,4-Dinitrophenol	ND	5.0	5.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2,4-Dinitrotoluene	ND	0.80	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2,6-Dinitrotoluene	ND	0.80	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Fluoranthene	ND	0.80	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Fluorene	ND	0.80	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C

Analyte	Result	MDL	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	ELAP#	Notes
G1271CUFS - 04 (21J0119-04) Water Sampled: 10/04/21 00:00 Received: 09/23/21 23:20												
Hexachlorobenzene	ND	0.90	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Hexachlorobutadiene	ND	0.80	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Hexachlorocyclopentadiene	ND	4.0	4.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Hexachloroethane	ND	0.60	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Indeno (1,2,3-cd) pyrene	ND	0.60	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Isophorone	ND	0.30	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2-Methylnaphthalene	ND	1.0	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2-Methylphenol (o-cresol)	ND	0.60	5.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
3 & 4-Methylphenol (m & p-cresol)	ND	0.60	5.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
N-Nitrosodi-n-propylamine	ND	0.80	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
N-Nitrosodimethylamine	ND	0.30	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
N-Nitrosodiphenylamine	ND	1.0	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Naphthalene	ND	0.70	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2-Nitroaniline	ND	1.0	20	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
3-Nitroaniline	ND	5.0	50	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
4-Nitroaniline	ND	4.0	20	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2-Nitrophenol	ND	0.90	10	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
4-Nitrophenol	ND	3.0	5.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Nitrobenzene	ND	0.90	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Pentachlorophenol	ND	4.0	5.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Phenanthrene	ND	0.90	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Phenol	ND	0.50	1.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Pyrene	ND	0.40	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
1,2,4-Trichlorobenzene	ND	0.60	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2,4,5-Trichlorophenol	ND	0.60	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
2,4,6-Trichlorophenol	ND	0.70	2.0	ug/L	1	AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	U
Surrogate: 2-Fluorobiphenyl		101 %	27-119			AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	
Surrogate: 2-Fluorophenol		47.8 %	7-85			AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	
Surrogate: Nitrobenzene-d5		89.2 %	15-314			AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	
Surrogate: p-Terphenyl-d14		111 %	36-141			AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	
Surrogate: Phenol-d6		34.9 %	1-65			AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	
Surrogate: 2,4,6-Tribromophenol		98.6 %	4-168			AJ13246	10/05/21 06:00	10/07/21 08:53	EPA 8270C/625	JV	1551	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC
 262 Rickenbacker Circle
 Livermore CA, 94551

Project Manager: Quality Control Manager
 Project: QC- 32oz Glass Boston Round Amber
 Project Number: Lot Number G1271CUFS

Reported:
 10/27/21 11:37

Benzidines by EPA Method 8270C

Analyte	Result	MDL	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	ELAP#	Notes
G1271CUFS - 04 (21J0119-04) Water Sampled: 10/04/21 00:00 Received: 09/23/21 23:20												
Benzidine	ND	3.0	5.0	ug/L	1	AJ13243	10/05/21 08:00	10/13/21 12:07	EPA 8270C	NBH	1551	U
3,3'-Dichlorobenzidine	ND	2.0	5.0	ug/L	1	AJ13243	10/05/21 08:00	10/13/21 12:07	EPA 8270C	NBH	1551	U
<i>Surrogate: 2-Fluorobiphenyl</i>		81.6 %		27-119		AJ13243	10/05/21 08:00	10/13/21 12:07	EPA 8270C	NBH	1551	
<i>Surrogate: Nitrobenzene-d5</i>		82.8 %		15-314		AJ13243	10/05/21 08:00	10/13/21 12:07	EPA 8270C	NBH	1551	
<i>Surrogate: p-Terphenyl-d14</i>		83.0 %		36-141		AJ13243	10/05/21 08:00	10/13/21 12:07	EPA 8270C	NBH	1551	



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC
 262 Rickenbacker Circle
 Livermore CA, 94551

Project Manager: Quality Control Manager
 Project: QC- 32oz Glass Boston Round Amber
 Project Number: Lot Number G1271CUFS

Reported:
 10/27/21 11:37

Semivolatile Organic Compounds by EPA Method 8270C SIM

Analyte	Result	MDL	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Analyst	ELAP#	Notes
G1271CUFS - 05 (21J0119-05) Water Sampled: 10/04/21 00:00 Received: 09/23/21 23:20												
Acenaphthylene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Anthracene	ND	0.090	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Benzo (a) anthracene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Benzo (a) pyrene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Benzo (b) fluoranthene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Benzo (g,h,i) perylene	ND	0.10	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Benzo (k) fluoranthene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Chrysene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Dibenz (a,h) anthracene	ND	0.10	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Fluorene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Indeno (1,2,3-cd) pyrene	ND	0.030	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Phenanthrene	ND	0.10	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Pyrene	ND	0.20	0.20	ug/L	1	AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	U
Surrogate: 2-Fluorobiphenyl		87.7 %	24-119			AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	
Surrogate: Nitrobenzene-d5		87.7 %	15-314			AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	
Surrogate: p-Terphenyl-d14		101 %	37-139			AJ13401	10/05/21 06:00	10/07/21 23:39	EPA 8270SIM	JV	1551	



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Organochlorine Pesticides and PCBs by EPA Method 608.3 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13545 - SVOAs in Water GC

Blank (AJ13545-BLK1)

Prepared: 10/11/21 Analyzed: 10/14/21

Aldrin	ND	0.0020	0.0050	ug/L							U
alpha-BHC	ND	0.0040	0.010	ug/L							U
beta-BHC	ND	0.0040	0.0050	ug/L							U
delta-BHC	ND	0.0020	0.0050	ug/L							U
gamma-BHC (Lindane)	ND	0.0040	0.010	ug/L							U
Chlordane (tech)	ND	0.010	0.050	ug/L							U
4,4'-DDD	ND	0.0050	0.020	ug/L							U
4,4'-DDE	ND	0.0040	0.020	ug/L							U
4,4'-DDT	ND	0.0030	0.010	ug/L							U
Dieldrin	ND	0.0050	0.010	ug/L							U
Endosulfan I	ND	0.0040	0.010	ug/L							U
Endosulfan II	ND	0.0050	0.010	ug/L							U
Endosulfan sulfate	ND	0.0030	0.050	ug/L							U
Endrin	ND	0.0080	0.010	ug/L							U
Endrin aldehyde	ND	0.0040	0.040	ug/L							U
Heptachlor	ND	0.0050	0.010	ug/L							U
Heptachlor epoxide	ND	0.0090	0.010	ug/L							U
Methoxychlor	ND	0.010	0.50	ug/L							U
Toxaphene	ND	0.20	0.50	ug/L							U
PCB-1016	ND	0.080	0.50	ug/L							U
PCB-1221	ND	0.20	0.50	ug/L							U
PCB-1232	ND	0.10	0.50	ug/L							U
PCB-1242	ND	0.040	0.50	ug/L							U
PCB-1248	ND	0.060	0.50	ug/L							U
PCB-1254	ND	0.040	0.50	ug/L							U
PCB-1260	ND	0.040	0.50	ug/L							U
Surrogate: Dibutylchlorodate	0.388			ug/L	0.422		92.0	10-147			



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Organochlorine Pesticides and PCBs by EPA Method 608.3 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13545 - SVOAs in Water GC

LCS (AJ13545-BS1)		Prepared: 10/11/21 Analyzed: 10/14/21									
Aldrin	0.173	0.0020	0.0050	ug/L	0.240		72.0	42-140			
alpha-BHC	0.195	0.0040	0.010	ug/L	0.240		81.4	37-140			
beta-BHC	0.224	0.0040	0.0050	ug/L	0.240		93.4	17-147			
delta-BHC	0.239	0.0020	0.0050	ug/L	0.240		99.5	19-140			
gamma-BHC (Lindane)	0.207	0.0040	0.010	ug/L	0.240		86.3	32-140			
4,4'-DDD	0.209	0.0050	0.020	ug/L	0.240		87.3	25-160			
4,4'-DDT	0.260	0.0030	0.010	ug/L	0.240		108	31-141			
4,4'-DDE	0.231	0.0040	0.020	ug/L	0.240		96.4	30-145			
Dieldrin	0.237	0.0050	0.010	ug/L	0.240		98.8	36-146			
Endosulfan I	0.232	0.0040	0.010	ug/L	0.240		96.9	45-153			
Endosulfan II	0.240	0.0050	0.010	ug/L	0.240		100	0-202			
Endosulfan sulfate	0.221	0.0030	0.050	ug/L	0.240		92.1	26-144			
Endrin	0.281	0.0080	0.010	ug/L	0.240		117	30-147			
Endrin aldehyde	0.225	0.0040	0.040	ug/L	0.240		93.7	45-140			
Heptachlor	0.197	0.0050	0.010	ug/L	0.240		82.1	34-140			
Heptachlor epoxide	0.230	0.0090	0.010	ug/L	0.240		95.7	37-142			
Methoxychlor	0.257	0.010	0.50	ug/L	0.240		107	45-140			J
Surrogate: Dibutylchloroendate	0.374			ug/L	0.422		88.5	10-147			

LCS (AJ13545-BS2)		Prepared: 10/11/21 Analyzed: 10/15/21									
PCB-1016	1.92	0.080	0.50	ug/L	2.00		96.0	50-140			
PCB-1260	2.07	0.040	0.50	ug/L	2.00		104	8-140			

Matrix Spike (AJ13545-MS1)		Source: 21J0942-01			Prepared: 10/11/21 Analyzed: 10/15/21						
Aldrin	0.181	0.0020	0.0050	ug/L	0.240	ND	75.6	42-140			
alpha-BHC	0.213	0.0040	0.010	ug/L	0.240	ND	88.6	37-140			
beta-BHC	0.274	0.0040	0.0050	ug/L	0.240	ND	114	17-147			
delta-BHC	0.264	0.0020	0.0050	ug/L	0.240	ND	110	19-140			
gamma-BHC (Lindane)	0.235	0.0040	0.010	ug/L	0.240	ND	98.0	32-140			
4,4'-DDD	0.228	0.0050	0.020	ug/L	0.240	ND	95.0	31-141			
4,4'-DDE	0.237	0.0040	0.020	ug/L	0.240	ND	98.7	30-145			
4,4'-DDT	0.238	0.0030	0.010	ug/L	0.240	ND	99.0	25-160			
Dieldrin	0.246	0.0050	0.010	ug/L	0.240	ND	103	36-146			
Endosulfan I	0.247	0.0040	0.010	ug/L	0.240	ND	103	45-153			
Endosulfan II	0.244	0.0050	0.010	ug/L	0.240	ND	102	2-202			
Endosulfan sulfate	0.239	0.0030	0.050	ug/L	0.240	ND	99.6	26-144			
Endrin	0.295	0.0080	0.010	ug/L	0.240	ND	123	30-147			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC
 262 Rickenbacker Circle
 Livermore CA, 94551

Project Manager: Quality Control Manager
 Project: QC- 32oz Glass Boston Round Amber
 Project Number: Lot Number G1271CUFS

Reported:
 10/27/21 11:37

Organochlorine Pesticides and PCBs by EPA Method 608.3 - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13545 - SVOAs in Water GC

Matrix Spike (AJ13545-MS1)		Source: 21J0942-01			Prepared: 10/11/21		Analyzed: 10/15/21	
Endrin aldehyde	0.248	0.0040	0.040	ug/L	0.240	ND	103	45-140
Heptachlor	0.218	0.0050	0.010	ug/L	0.240	ND	90.8	34-140
Heptachlor epoxide	0.250	0.0090	0.010	ug/L	0.240	ND	104	37-142
Methoxychlor	0.236	0.010	0.50	ug/L	0.240	ND	98.5	45-140
Surrogate: Dibutylchlorendate	0.362			ug/L	0.422		85.9	10-147

Matrix Spike Dup (AJ13545-MSD1)		Source: 21J0942-01			Prepared: 10/11/21		Analyzed: 10/15/21			
Aldrin	0.190	0.0020	0.0050	ug/L	0.240	ND	79.3	42-140	4.77	35
alpha-BHC	0.225	0.0040	0.010	ug/L	0.240	ND	93.6	37-140	5.52	36
beta-BHC	0.275	0.0040	0.0050	ug/L	0.240	ND	114	17-147	0.128	44
delta-BHC	0.264	0.0020	0.0050	ug/L	0.240	ND	110	19-140	0.176	52
gamma-BHC (Lindane)	0.242	0.0040	0.010	ug/L	0.240	ND	101	32-140	2.68	39
4,4'-DDD	0.230	0.0050	0.020	ug/L	0.240	ND	95.9	31-141	0.942	39
4,4'-DDE	0.248	0.0040	0.020	ug/L	0.240	ND	103	30-145	4.44	35
4,4'-DDT	0.247	0.0030	0.010	ug/L	0.240	ND	103	25-160	3.92	42
Dieldrin	0.253	0.0050	0.010	ug/L	0.240	ND	105	36-146	2.65	49
Endosulfan I	0.251	0.0040	0.010	ug/L	0.240	ND	104	45-153	1.37	28
Endosulfan II	0.250	0.0050	0.010	ug/L	0.240	ND	104	2-202	2.33	53
Endosulfan sulfate	0.251	0.0030	0.050	ug/L	0.240	ND	105	26-144	4.91	38
Endrin	0.293	0.0080	0.010	ug/L	0.240	ND	122	30-147	0.630	48
Endrin aldehyde	0.243	0.0040	0.040	ug/L	0.240	ND	101	45-140	2.12	35
Heptachlor	0.219	0.0050	0.010	ug/L	0.240	ND	91.1	34-140	0.339	43
Heptachlor epoxide	0.252	0.0090	0.010	ug/L	0.240	ND	105	37-142	0.806	36
Methoxychlor	0.246	0.010	0.50	ug/L	0.240	ND	103	45-140	4.13	35
Surrogate: Dibutylchlorendate	0.372			ug/L	0.422		88.2	10-147		

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Chlorinated Herbicides by EPA Method 8151A - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13382 - Herbicides

Blank (AJ13382-BLK1)		Prepared: 10/06/21 Analyzed: 10/08/21									
2,4-D	ND	1.0	1.0	ug/L							U
2,4-DB	ND	1.0	4.0	ug/L							U
2,4,5-T	ND	0.20	0.50	ug/L							U
2,4,5-TP (Silvex)	ND	0.20	0.50	ug/L							U
Dalapon	ND	2.0	6.0	ug/L							U
Dicamba	ND	0.10	0.40	ug/L							U
Dinoseb	ND	0.20	1.0	ug/L							U
Dichlorprop	ND	0.40	1.0	ug/L							U
Pentachlorophenol	ND	0.20	0.20	ug/L							U
Surrogate: DCAA	10.9			ug/L	11.4		95.8	62-124			

Matrix Spike (AJ13382-MS1)		Source: 2113742-03		Prepared: 10/06/21 Analyzed: 10/08/21							
2,4-D	10.1	1.0	1.0	ug/L	10.6	ND	95.6	70-130			
2,4-DB	14.0	1.0	4.0	ug/L	14.1	ND	99.2	70-130			
2,4,5-T	1.67	0.20	0.50	ug/L	1.76	ND	95.0	70-130			
2,4,5-TP (Silvex)	1.67	0.20	0.50	ug/L	1.76	ND	94.9	70-130			
Dalapon	22.9	2.0	6.0	ug/L	22.9	ND	100	70-130			
Dicamba	1.74	0.10	0.40	ug/L	1.76	ND	98.7	70-130			
Dinoseb	3.29	0.20	1.0	ug/L	3.52	ND	93.4	70-130			
Dichlorprop	5.16	0.40	1.0	ug/L	5.28	ND	97.8	70-130			
Pentachlorophenol	1.19	0.20	0.20	ug/L	1.32	ND	89.8	70-130			
Surrogate: DCAA	12.4			ug/L	11.4		108	62-124			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Organophosphorus Compounds by EPA Method 8141A - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ14235 - SVOAs in Water GC

Blank (AJ14235-BLK1)

Prepared: 10/20/21 Analyzed: 10/25/21

Azinphos ethyl	ND	1.0	2.0	ug/L							U
Azinphos methyl	ND	0.40	2.0	ug/L							U
Bolstar	ND	0.30	2.0	ug/L							U
Chlorpyrifos	ND	0.70	1.0	ug/L							U
Coumaphos	ND	0.80	2.0	ug/L							U
Demeton-o	ND	0.20	2.0	ug/L							U
Demeton-s	ND	0.30	2.0	ug/L							U
Diazinon	ND	0.30	0.50	ug/L							U
Dichlorvos	ND	0.40	2.0	ug/L							U
Dimethoate	ND	1.0	2.0	ug/L							U
Disulfoton	ND	0.30	2.0	ug/L							U
EPN	ND	0.50	2.0	ug/L							U
Ethion	ND	0.80	2.0	ug/L							U
Ethoprop (Ethoprophos)	ND	0.40	2.0	ug/L							U
Fensulfothion	ND	0.80	2.0	ug/L							U
Fenthion	ND	0.50	2.0	ug/L							U
Malathion	ND	0.80	2.0	ug/L							U
Mevinphos	ND	0.80	2.0	ug/L							U
Parathion	ND	0.90	2.0	ug/L							U
Parathion-methyl	ND	0.40	2.0	ug/L							U
Phorate	ND	0.40	2.0	ug/L							U
Ronnel	ND	0.30	2.0	ug/L							U
Stirofos	ND	0.30	2.0	ug/L							U
Thionazin	ND	1.0	2.0	ug/L							U
Tokuthion (Prothiofos)	ND	0.40	2.0	ug/L							U
Trichloronate	ND	0.40	2.0	ug/L							U
Surrogate: Tributyl phosphate	1.82			ug/L	2.00		91.2	37-115			
Surrogate: Triphenyl phosphate	2.05			ug/L	2.00		103	44-145			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Organophosphorus Compounds by EPA Method 8141A - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ14235 - SVOAs in Water GC

LCS (AJ14235-BS1)		Prepared: 10/20/21 Analyzed: 10/26/21									
Azinphos ethyl	2.54	1.0	2.0	ug/L	2.00		127	44-154			
Azinphos methyl	2.46	0.40	2.0	ug/L	2.00		123	42-147			
Bolstar	2.52	0.30	2.0	ug/L	2.00		126	35-144			
Chlorpyrifos	2.45	0.70	1.0	ug/L	2.00		122	50-134			
Coumaphos	2.37	0.80	2.0	ug/L	2.00		119	47-152			
Demeton-o	0.781	0.20	2.0	ug/L	0.750		104	18-128			J
Demeton-s	1.22	0.30	2.0	ug/L	1.25		97.9	43-125			J
Diazinon	2.31	0.30	0.50	ug/L	2.00		116	50-121			
Dichlorvos	2.43	0.40	2.0	ug/L	2.00		121	77-143			
Dimethoate	2.20	1.0	2.0	ug/L	2.00		110	37-133			
Disulfoton	2.30	0.30	2.0	ug/L	2.00		115	50-119			
EPN	2.52	0.50	2.0	ug/L	2.00		126	55-140			
Ethion	2.39	0.80	2.0	ug/L	2.00		120	38-137			
Ethoprop (Ethoprophos)	2.40	0.40	2.0	ug/L	2.00		120	67-130			
Fensulfothion	2.32	0.80	2.0	ug/L	2.00		116	35-157			
Fenthion	2.32	0.50	2.0	ug/L	2.00		116	56-125			
Malathion	2.11	0.80	2.0	ug/L	2.00		106	39-127			
Mevinphos	2.44	0.80	2.0	ug/L	2.00		122	58-125			
Parathion	2.21	0.90	2.0	ug/L	2.00		111	33-137			
Parathion-methyl	2.25	0.40	2.0	ug/L	2.00		113	52-134			
Phorate	2.28	0.40	2.0	ug/L	2.00		114	58-119			
Ronnel	2.16	0.30	2.0	ug/L	2.00		108	41-135			
Stirofos	2.29	0.30	2.0	ug/L	2.00		115	34-151			
Thionazin	2.33	1.0	2.0	ug/L	2.00		116	60-118			
Tokuthion (Prothiofos)	2.33	0.40	2.0	ug/L	2.00		116	41-136			
Trichloronate	2.15	0.40	2.0	ug/L	2.00		107	36-146			
Surrogate: Tributyl phosphate	1.91			ug/L	2.00		95.5	37-115			
Surrogate: Triphenyl phosphate	1.88			ug/L	2.00		93.8	44-145			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Organophosphorus Compounds by EPA Method 8141A - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ14235 - SVOAs in Water GC

LCS Dup (AJ14235-BSD1)

Prepared: 10/20/21 Analyzed: 10/26/21

Azinphos ethyl	2.36	1.0	2.0	ug/L	2.00	118	44-154	7.59	30	
Azinphos methyl	2.40	0.40	2.0	ug/L	2.00	120	42-147	2.77	30	
Bolstar	2.42	0.30	2.0	ug/L	2.00	121	35-144	3.88	30	
Chlorpyrifos	2.51	0.70	1.0	ug/L	2.00	126	50-134	2.63	30	
Coumaphos	2.26	0.80	2.0	ug/L	2.00	113	47-152	5.09	30	
Demeton-o	0.662	0.20	2.0	ug/L	0.750	88.3	18-128	16.4	30	J
Demeton-s	1.25	0.30	2.0	ug/L	1.25	100	43-125	2.39	30	J
Diazinon	2.24	0.30	0.50	ug/L	2.00	112	50-121	3.27	30	
Dichlorvos	2.14	0.40	2.0	ug/L	2.00	107	77-143	12.4	30	
Dimethoate	2.15	1.0	2.0	ug/L	2.00	108	37-133	2.18	30	
Disulfoton	2.20	0.30	2.0	ug/L	2.00	110	50-119	4.55	30	
EPN	2.40	0.50	2.0	ug/L	2.00	120	55-140	4.96	30	
Ethion	2.31	0.80	2.0	ug/L	2.00	115	38-137	3.59	30	
Ethoprop (Ethoprophos)	2.28	0.40	2.0	ug/L	2.00	114	67-130	4.92	30	
Fensulfothion	2.53	0.80	2.0	ug/L	2.00	127	35-157	8.75	30	
Fenthion	2.23	0.50	2.0	ug/L	2.00	111	56-125	4.09	30	
Malathion	2.16	0.80	2.0	ug/L	2.00	108	39-127	2.08	30	
Mevinphos	2.43	0.80	2.0	ug/L	2.00	121	58-125	0.599	30	
Parathion	2.29	0.90	2.0	ug/L	2.00	115	33-137	3.58	30	
Parathion-methyl	2.23	0.40	2.0	ug/L	2.00	112	52-134	0.984	30	
Phorate	2.23	0.40	2.0	ug/L	2.00	111	58-119	2.40	30	
Ronnel	2.31	0.30	2.0	ug/L	2.00	116	41-135	6.89	30	
Stirofos	2.19	0.30	2.0	ug/L	2.00	110	34-151	4.46	30	
Thionazin	2.04	1.0	2.0	ug/L	2.00	102	60-118	13.2	30	
Tokuthion (Prothiofos)	2.26	0.40	2.0	ug/L	2.00	113	41-136	2.88	30	
Trichloronate	2.07	0.40	2.0	ug/L	2.00	104	36-146	3.40	30	
Surrogate: Tributyl phosphate	1.73			ug/L	2.00	86.4	37-115			
Surrogate: Triphenyl phosphate	1.76			ug/L	2.00	87.8	44-145			



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatle Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

Blank (AJ13246-BLK1)

Prepared: 10/05/21 Analyzed: 10/07/21

Acenaphthene	ND	0.60	1.0	ug/L							U
Acenaphthylene	ND	0.50	2.0	ug/L							U
Anthracene	ND	0.40	2.0	ug/L							U
Benzo (a) anthracene	ND	0.40	2.0	ug/L							U
Benzo (a) pyrene	ND	0.50	2.0	ug/L							U
Benzo (b) fluoranthene	ND	0.60	2.0	ug/L							U
Benzo (g,h,i) perylene	ND	0.90	2.0	ug/L							U
Benzo (k) fluoranthene	ND	0.50	2.0	ug/L							U
Bis(2-chloroethoxy)methane	ND	0.90	2.0	ug/L							U
Bis(2-chloroethyl)ether	ND	0.90	1.0	ug/L							U
Bis(2-chloroisopropyl)ether	ND	1.0	1.0	ug/L							U
Bis(2-ethylhexyl)phthalate	ND	3.0	5.0	ug/L							U
4-Bromophenyl phenyl ether	ND	1.0	4.0	ug/L							U
Butyl benzyl phthalate	ND	0.60	5.0	ug/L							U
4-Chloro-3-methylphenol	ND	1.0	1.0	ug/L							U
4-Chloroaniline	ND	1.0	14	ug/L							U
2-Chloronaphthalene	ND	0.60	2.0	ug/L							U
2-Chlorophenol	ND	0.70	2.0	ug/L							U
4-Chlorophenyl phenyl ether	ND	0.90	4.0	ug/L							U
Chrysene	ND	0.80	2.0	ug/L							U
Di-n-butyl phthalate	ND	0.90	10	ug/L							U
Di-n-octyl phthalate	ND	0.60	10	ug/L							U
Dibenz (a,h) anthracene	ND	0.80	2.0	ug/L							U
Dibenzofuran	ND	0.90	4.0	ug/L							U
1,2-Dichlorobenzene	ND	0.60	1.0	ug/L							U
1,3-Dichlorobenzene	ND	0.60	1.0	ug/L							U
1,4-Dichlorobenzene	ND	0.60	1.0	ug/L							U
2,4-Dichlorophenol	ND	0.70	2.0	ug/L							U
Diethyl phthalate	ND	0.30	2.0	ug/L							U
Dimethyl phthalate	ND	0.80	2.0	ug/L							U
2,4-Dimethylphenol	ND	1.0	2.0	ug/L							U
4,6-Dinitro-2-methylphenol	ND	3.0	4.0	ug/L							U
2,4-Dinitrophenol	ND	5.0	5.0	ug/L							U
2,4-Dinitrotoluene	ND	0.80	2.0	ug/L							U
2,6-Dinitrotoluene	ND	0.80	2.0	ug/L							U
Fluoranthene	ND	0.80	1.0	ug/L							U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

Blank (AJ13246-BLK1)

Prepared: 10/05/21 Analyzed: 10/07/21

Fluorene	ND	0.80	2.0	ug/L							U
Hexachlorobenzene	ND	0.90	1.0	ug/L							U
Hexachlorobutadiene	ND	0.80	1.0	ug/L							U
Hexachlorocyclopentadiene	ND	4.0	4.0	ug/L							U
Hexachloroethane	ND	0.60	1.0	ug/L							U
Indeno (1,2,3-cd) pyrene	ND	0.60	2.0	ug/L							U
Isophorone	ND	0.30	1.0	ug/L							U
2-Methylnaphthalene	ND	1.0	2.0	ug/L							U
2-Methylphenol (o-cresol)	ND	0.60	5.0	ug/L							U
3 & 4-Methylphenol (m & p-cresol)	ND	0.60	5.0	ug/L							U
N-Nitrosodi-n-propylamine	ND	0.80	2.0	ug/L							U
N-Nitrosodimethylamine	ND	0.30	2.0	ug/L							U
N-Nitrosodiphenylamine	ND	1.0	1.0	ug/L							U
Naphthalene	ND	0.70	1.0	ug/L							U
2-Nitroaniline	ND	1.0	20	ug/L							U
3-Nitroaniline	ND	5.0	50	ug/L							U
4-Nitroaniline	ND	4.0	20	ug/L							U
2-Nitrophenol	ND	0.90	10	ug/L							U
4-Nitrophenol	ND	3.0	5.0	ug/L							U
Nitrobenzene	ND	0.90	1.0	ug/L							U
Pentachlorophenol	ND	4.0	5.0	ug/L							U
Phenanthrene	ND	0.90	2.0	ug/L							U
Phenol	ND	0.50	1.0	ug/L							U
Pyrene	ND	0.40	2.0	ug/L							U
1,2,4-Trichlorobenzene	ND	0.60	2.0	ug/L							U
2,4,5-Trichlorophenol	ND	0.60	2.0	ug/L							U
2,4,6-Trichlorophenol	ND	0.70	2.0	ug/L							U
Surrogate: 2-Fluorobiphenyl	40.7			ug/L	40.0		102	27-119			
Surrogate: 2-Fluorophenol	19.0			ug/L	40.0		47.6	7-85			
Surrogate: Nitrobenzene-d5	36.4			ug/L	40.0		90.9	15-314			
Surrogate: p-Terphenyl-d14	41.6			ug/L	40.0		104	36-141			
Surrogate: Phenol-d6	12.6			ug/L	40.0		31.5	1-65			
Surrogate: 2,4,6-Tribromophenol	28.9			ug/L	40.0		72.3	4-168			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

LCS (AJ13246-BS1)

Prepared: 10/05/21 Analyzed: 10/06/21

Acenaphthene	39	0.60	1.0	ug/L	40.0		96.3	60-132			
Acenaphthylene	39	0.50	2.0	ug/L	40.0		98.1	54-126			
Anthracene	40	0.40	2.0	ug/L	40.0		100	43-120			
Benzo (a) anthracene	41	0.40	2.0	ug/L	40.0		102	42-133			
Benzo (a) pyrene	41	0.50	2.0	ug/L	40.0		103	32-148			
Benzo (b) fluoranthene	42	0.60	2.0	ug/L	40.0		104	42-140			
Benzo (g,h,i) perylene	41	0.90	2.0	ug/L	40.0		103	2-195			
Benzo (k) fluoranthene	43	0.50	2.0	ug/L	40.0		108	25-146			
Bis(2-chloroethoxy)methane	38	0.90	2.0	ug/L	40.0		95.8	49-165			
Bis(2-chloroethyl)ether	32	0.90	1.0	ug/L	40.0		80.4	43-126			
Bis(2-chloroisopropyl)ether	34	1.0	1.0	ug/L	40.0		85.3	63-139			
Bis(2-ethylhexyl)phthalate	43	3.0	5.0	ug/L	40.0		106	29-137			
4-Bromophenyl phenyl ether	41	1.0	4.0	ug/L	40.0		102	65-120			
Butyl benzyl phthalate	31	0.60	5.0	ug/L	40.0		77.0	2-140			
4-Chloro-3-methylphenol	41	1.0	1.0	ug/L	40.0		101	41-128			
4-Chloroaniline	59	1.0	14	ug/L	60.0		99.0	38-114			
2-Chloronaphthalene	39	0.60	2.0	ug/L	40.0		98.0	65-120			
2-Chlorophenol	32	0.70	2.0	ug/L	40.0		81.2	36-120			
4-Chlorophenyl phenyl ether	41	0.90	4.0	ug/L	40.0		102	38-145			
Chrysene	43	0.80	2.0	ug/L	40.0		107	44-140			
Di-n-butyl phthalate	46	0.90	10	ug/L	40.0		115	8-120			
Di-n-octyl phthalate	41	0.60	10	ug/L	40.0		104	19-132			
Dibenz (a,h) anthracene	42	0.80	2.0	ug/L	40.0		104	2-200			
Dibenzofuran	39	0.90	4.0	ug/L	40.0		98.7	57-107			
1,2-Dichlorobenzene	33	0.60	1.0	ug/L	40.0		81.6	33-93			
1,3-Dichlorobenzene	31	0.60	1.0	ug/L	40.0		77.7	31-92			
1,4-Dichlorobenzene	32	0.60	1.0	ug/L	40.0		79.1	32-91			
2,4-Dichlorophenol	37	0.70	2.0	ug/L	40.0		92.4	53-122			
Diethyl phthalate	31	0.30	2.0	ug/L	40.0		77.9	2-120			
Dimethyl phthalate	21	0.80	2.0	ug/L	40.0		53.0	2-120			
2,4-Dimethylphenol	42	1.0	2.0	ug/L	50.0		83.3	42-120			
4,6-Dinitro-2-methylphenol	32	3.0	4.0	ug/L	40.0		80.7	53-130			
2,4-Dinitrophenol	34	5.0	5.0	ug/L	40.0		85.7	10-173			
2,4-Dinitrotoluene	41	0.80	2.0	ug/L	40.0		102	48-127			
2,6-Dinitrotoluene	42	0.80	2.0	ug/L	40.0		105	68-137			
Fluoranthene	40	0.80	1.0	ug/L	40.0		101	43-121			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

LCS (AJ13246-BS1)		Prepared: 10/05/21 Analyzed: 10/06/21									
Fluorene	40	0.80	2.0	ug/L	40.0		99.3	70-120			
Hexachlorobenzene	40	0.90	1.0	ug/L	40.0		99.8	8-142			
Hexachlorobutadiene	36	0.80	1.0	ug/L	40.0		89.8	38-120			
Hexachlorocyclopentadiene	31	4.0	4.0	ug/L	40.0		78.2	16-93			
Hexachloroethane	32	0.60	1.0	ug/L	40.0		78.9	55-120			
Indeno (1,2,3-cd) pyrene	43	0.60	2.0	ug/L	40.0		108	2-151			
Isophorone	37	0.30	1.0	ug/L	40.0		93.0	47-180			
2-Methylnaphthalene	41	1.0	2.0	ug/L	40.0		103	45-106			
2-Methylphenol (o-cresol)	33	0.60	5.0	ug/L	40.0		81.6	36-102			
3 & 4-Methylphenol (m & p-cresol)	31	0.60	5.0	ug/L	40.0		77.4	38-93			
N-Nitrosodi-n-propylamine	36	0.80	2.0	ug/L	40.0		89.3	14-198			
N-Nitrosodimethylamine	23	0.30	2.0	ug/L	40.0		57.8	19-81			
N-Nitrosodiphenylamine	31	1.0	1.0	ug/L	40.0		78.0	54-111			
Naphthalene	38	0.70	1.0	ug/L	40.0		94.8	36-120			
2-Nitroaniline	59	1.0	20	ug/L	60.0		99.0	53-128			
3-Nitroaniline	59	5.0	50	ug/L	60.0		98.0	50-137			
4-Nitroaniline	56	4.0	20	ug/L	60.0		93.9	25-191			
2-Nitrophenol	35	0.90	10	ug/L	40.0		86.4	45-167			
4-Nitrophenol	21	3.0	5.0	ug/L	40.0		53.0	13-129			
Nitrobenzene	37	0.90	1.0	ug/L	40.0		92.2	54-158			
Pentachlorophenol	37	4.0	5.0	ug/L	40.0		93.3	38-152			
Phenanthrene	41	0.90	2.0	ug/L	40.0		101	65-120			
Phenol	19	0.50	1.0	ug/L	40.0		46.5	17-120			
Pyrene	41	0.40	2.0	ug/L	40.0		103	70-120			
1,2,4-Trichlorobenzene	36	0.60	2.0	ug/L	40.0		90.6	57-130			
2,4,5-Trichlorophenol	38	0.60	2.0	ug/L	40.0		95.4	44-141			
2,4,6-Trichlorophenol	36	0.70	2.0	ug/L	40.0		90.0	52-129			
Surrogate: 2-Fluorobiphenyl	39.8			ug/L	40.0		99.6	27-119			
Surrogate: 2-Fluorophenol	23.4			ug/L	40.0		58.4	7-85			
Surrogate: Nitrobenzene-d5	38.3			ug/L	40.0		95.8	15-314			
Surrogate: p-Terphenyl-d14	42.1			ug/L	40.0		105	36-141			
Surrogate: Phenol-d6	17.2			ug/L	40.0		43.0	1-65			
Surrogate: 2,4,6-Tribromophenol	36.4			ug/L	40.0		90.9	4-168			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

Matrix Spike (AJ13246-MS1)	Source: 2113566-01			Prepared: 10/05/21		Analyzed: 10/07/21		
Acenaphthene	36	0.60	1.0	ug/L	40.0	ND	90.0	47-145
Acenaphthylene	36	0.50	2.0	ug/L	40.0	ND	90.7	33-145
Anthracene	39	0.40	2.0	ug/L	40.0	ND	98.1	27-133
Benzo (a) anthracene	34	0.40	2.0	ug/L	40.0	ND	85.5	33-143
Benzo (a) pyrene	26	0.50	2.0	ug/L	40.0	ND	64.6	17-163
Benzo (b) fluoranthene	26	0.60	2.0	ug/L	40.0	ND	65.6	24-159
Benzo (g,h,i) perylene	16	0.90	2.0	ug/L	40.0	ND	40.9	2-219
Benzo (k) fluoranthene	27	0.50	2.0	ug/L	40.0	ND	67.6	11-162
Bis(2-chloroethoxy)methane	36	0.90	2.0	ug/L	40.0	ND	89.1	33-184
Bis(2-chloroethyl)ether	33	0.90	1.0	ug/L	40.0	ND	81.9	12-158
Bis(2-chloroisopropyl)ether	33	1.0	1.0	ug/L	40.0	ND	82.4	36-166
Bis(2-ethylhexyl)phthalate	15	3.0	5.0	ug/L	40.0	ND	38.6	8-158
4-Bromophenyl phenyl ether	38	1.0	4.0	ug/L	40.0	ND	95.8	53-127
Butyl benzyl phthalate	27	0.60	5.0	ug/L	40.0	ND	68.6	2-152
4-Chloro-3-methylphenol	36	1.0	1.0	ug/L	40.0	ND	91.0	22-147
4-Chloroaniline	56	1.0	14	ug/L	60.0	ND	93.8	9-1114
2-Chloronaphthalene	36	0.60	2.0	ug/L	40.0	ND	90.0	60-120
2-Chlorophenol	31	0.70	2.0	ug/L	40.0	ND	78.0	23-134
4-Chlorophenyl phenyl ether	37	0.90	4.0	ug/L	40.0	ND	92.7	25-158
Chrysene	35	0.80	2.0	ug/L	40.0	ND	88.5	17-168
Di-n-butyl phthalate	50	0.90	10	ug/L	40.0	12	95.7	1-120
Di-n-octyl phthalate	14	0.60	10	ug/L	40.0	ND	35.6	4-146
Dibenz (a,h) anthracene	17	0.80	2.0	ug/L	40.0	ND	41.6	2-227
Dibenzofuran	37	0.90	4.0	ug/L	40.0	ND	93.1	57-107
1,2-Dichlorobenzene	32	0.60	1.0	ug/L	40.0	ND	79.6	7-104
1,3-Dichlorobenzene	30	0.60	1.0	ug/L	40.0	ND	75.8	1-106
1,4-Dichlorobenzene	31	0.60	1.0	ug/L	40.0	ND	77.0	3-105
2,4-Dichlorophenol	33	0.70	2.0	ug/L	40.0	ND	83.5	39-135
Diethyl phthalate	29	0.30	2.0	ug/L	40.0	ND	73.2	2-120
Dimethyl phthalate	21	0.80	2.0	ug/L	40.0	ND	51.4	2-120
2,4-Dimethylphenol	27	1.0	2.0	ug/L	50.0	ND	54.9	32-120
4,6-Dinitro-2-methylphenol	15	3.0	4.0	ug/L	40.0	ND	38.2	7-181
2,4-Dinitrophenol	6.7	5.0	5.0	ug/L	40.0	ND	16.8	10-191
2,4-Dinitrotoluene	41	0.80	2.0	ug/L	40.0	ND	103	39-139
2,6-Dinitrotoluene	40	0.80	2.0	ug/L	40.0	ND	101	50-158
Fluoranthene	39	0.80	1.0	ug/L	40.0	ND	97.7	26-137

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

Matrix Spike (AJ13246-MS1)	Source: 2113566-01	Prepared: 10/05/21	Analyzed: 10/07/21				
Fluorene	37	0.80	2.0 ug/L	40.0	ND	92.5	59-121
Hexachlorobenzene	38	0.90	1.0 ug/L	40.0	ND	94.0	2-152
Hexachlorobutadiene	30	0.80	1.0 ug/L	40.0	ND	75.8	24-120
Hexachlorocyclopentadiene	28	4.0	4.0 ug/L	40.0	ND	69.4	10-93
Hexachloroethane	30	0.60	1.0 ug/L	40.0	ND	74.2	40-120
Indeno (1,2,3-cd) pyrene	17	0.60	2.0 ug/L	40.0	ND	43.7	2-171
Isophorone	35	0.30	1.0 ug/L	40.0	ND	86.9	21-196
2-Methylnaphthalene	36	1.0	2.0 ug/L	40.0	ND	91.0	44-106
2-Methylphenol (o-cresol)	30	0.60	5.0 ug/L	40.0	ND	75.9	9-124
3 & 4-Methylphenol (m & p-cresol)	29	0.60	5.0 ug/L	40.0	ND	73.2	12-99
N-Nitrosodi-n-propylamine	36	0.80	2.0 ug/L	40.0	ND	89.8	2-230
N-Nitrosodimethylamine	22	0.30	2.0 ug/L	40.0	ND	54.2	10-84
N-Nitrosodiphenylamine	40	1.0	1.0 ug/L	40.0	ND	101	40-115
Naphthalene	34	0.70	1.0 ug/L	40.0	ND	84.2	21-133
2-Nitroaniline	59	1.0	20 ug/L	60.0	ND	98.3	48-128
3-Nitroaniline	56	5.0	50 ug/L	60.0	ND	94.1	17-137
4-Nitroaniline	55	4.0	20 ug/L	60.0	ND	91.7	13-191
2-Nitrophenol	33	0.90	10 ug/L	40.0	ND	82.6	29-182
4-Nitrophenol	7.7	3.0	5.0 ug/L	40.0	ND	19.2	7-132
Nitrobenzene	34	0.90	1.0 ug/L	40.0	ND	84.8	35-180
Pentachlorophenol	ND	4.0	5.0 ug/L	40.0	ND		14-176
Phenanthrene	39	0.90	2.0 ug/L	40.0	ND	97.3	54-120
Phenol	17	0.50	1.0 ug/L	40.0	ND	42.9	5-120
Pyrene	39	0.40	2.0 ug/L	40.0	ND	97.6	52-120
1,2,4-Trichlorobenzene	32	0.60	2.0 ug/L	40.0	ND	80.4	44-142
2,4,5-Trichlorophenol	32	0.60	2.0 ug/L	40.0	ND	79.7	65-125
2,4,6-Trichlorophenol	26	0.70	2.0 ug/L	40.0	ND	65.5	37-144
Surrogate: 2-Fluorobiphenyl	36.7		ug/L	40.0		91.8	27-119
Surrogate: 2-Fluorophenol	18.9		ug/L	40.0		47.3	7-85
Surrogate: Nitrobenzene-d5	35.7		ug/L	40.0		89.2	15-314
Surrogate: p-Terphenyl-d14	34.0		ug/L	40.0		84.9	36-141
Surrogate: Phenol-d6	16.0		ug/L	40.0		39.9	1-65
Surrogate: 2,4,6-Tribromophenol	29.1		ug/L	40.0		72.8	4-168

QM-05, U

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatle Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

Matrix Spike Dup (AJ13246-MSD1)	Source: 2113566-01		Prepared: 10/05/21		Analyzed: 10/07/21						
Acenaphthene	35	0.60	1.0	ug/L	40.0	ND	88.2	47-145	2.10	48	
Acenaphthylene	36	0.50	2.0	ug/L	40.0	ND	89.6	33-145	1.19	74	
Anthracene	39	0.40	2.0	ug/L	40.0	ND	96.5	27-133	1.64	66	
Benzo (a) anthracene	32	0.40	2.0	ug/L	40.0	ND	81.0	33-143	5.41	53	
Benzo (a) pyrene	29	0.50	2.0	ug/L	40.0	ND	71.5	17-163	10.1	72	
Benzo (b) fluoranthene	29	0.60	2.0	ug/L	40.0	ND	73.0	24-159	10.8	71	
Benzo (g,h,i) perylene	20	0.90	2.0	ug/L	40.0	ND	49.7	2-219	19.5	97	
Benzo (k) fluoranthene	29	0.50	2.0	ug/L	40.0	ND	72.9	11-162	7.58	63	
Bis(2-chloroethoxy)methane	35	0.90	2.0	ug/L	40.0	ND	86.4	33-184	3.10	54	
Bis(2-chloroethyl)ether	31	0.90	1.0	ug/L	40.0	ND	77.0	12-158	6.20	108	
Bis(2-chloroisopropyl)ether	31	1.0	1.0	ug/L	40.0	ND	76.6	36-166	7.30	76	
Bis(2-ethylhexyl)phthalate	19	3.0	5.0	ug/L	40.0	ND	46.5	8-158	18.4	82	
4-Bromophenyl phenyl ether	37	1.0	4.0	ug/L	40.0	ND	93.4	53-127	2.56	43	
Butyl benzyl phthalate	25	0.60	5.0	ug/L	40.0	ND	63.3	2-152	8.04	60	
4-Chloro-3-methylphenol	35	1.0	1.0	ug/L	40.0	ND	88.6	22-147	2.67	73	
4-Chloroaniline	58	1.0	14	ug/L	60.0	ND	97.0	9-1114	3.39	25	
2-Chloronaphthalene	35	0.60	2.0	ug/L	40.0	ND	87.1	60-120	3.33	24	
2-Chlorophenol	29	0.70	2.0	ug/L	40.0	ND	73.0	23-134	6.75	61	
4-Chlorophenyl phenyl ether	37	0.90	4.0	ug/L	40.0	ND	93.7	25-158	1.07	61	
Chrysene	35	0.80	2.0	ug/L	40.0	ND	87.2	17-168	1.54	87	
Di-n-butyl phthalate	40	0.90	10	ug/L	40.0	12	70.4	1-120	22.5	47	
Di-n-octyl phthalate	18	0.60	10	ug/L	40.0	ND	45.4	4-146	24.2	69	
Dibenz (a,h) anthracene	20	0.80	2.0	ug/L	40.0	ND	50.8	2-227	20.0	126	
Dibenzofuran	37	0.90	4.0	ug/L	40.0	ND	91.6	57-107	1.65	29	
1,2-Dichlorobenzene	30	0.60	1.0	ug/L	40.0	ND	74.2	7-104	6.99	41	
1,3-Dichlorobenzene	29	0.60	1.0	ug/L	40.0	ND	71.8	1-106	5.42	41	
1,4-Dichlorobenzene	29	0.60	1.0	ug/L	40.0	ND	73.6	3-105	4.48	42	
2,4-Dichlorophenol	33	0.70	2.0	ug/L	40.0	ND	81.6	39-135	2.27	50	
Diethyl phthalate	28	0.30	2.0	ug/L	40.0	ND	70.1	2-120	4.26	100	
Dimethyl phthalate	20	0.80	2.0	ug/L	40.0	ND	49.0	2-120	4.78	183	
2,4-Dimethylphenol	22	1.0	2.0	ug/L	50.0	ND	44.1	32-120	21.7	58	
4,6-Dinitro-2-methylphenol	10	3.0	4.0	ug/L	40.0	ND	25.4	7-181	39.9	203	
2,4-Dinitrophenol	ND	5.0	5.0	ug/L	40.0	ND		10-191	200	132	QM-05, U
2,4-Dinitrotoluene	42	0.80	2.0	ug/L	40.0	ND	104	39-139	1.62	42	
2,6-Dinitrotoluene	41	0.80	2.0	ug/L	40.0	ND	102	50-158	1.30	48	
Fluoranthene	40	0.80	1.0	ug/L	40.0	ND	101	26-137	3.07	66	

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13246 - SVOAs in Water GCMS

Matrix Spike Dup (AJ13246-MSD1)	Source: 2113566-01	Prepared: 10/05/21	Analyzed: 10/07/21							
Fluorene	37	0.80	2.0 ug/L	40.0	ND	91.6	59-121	0.951	38	
Hexachlorobenzene	36	0.90	1.0 ug/L	40.0	ND	90.6	2-152	3.66	55	
Hexachlorobutadiene	31	0.80	1.0 ug/L	40.0	ND	76.6	24-120	0.951	62	
Hexachlorocyclopentadiene	23	4.0	4.0 ug/L	40.0	ND	58.1	10-93	17.8	55	
Hexachloroethane	28	0.60	1.0 ug/L	40.0	ND	70.6	40-120	4.94	52	
Indeno (1,2,3-cd) pyrene	21	0.60	2.0 ug/L	40.0	ND	53.1	2-171	19.4	99	
Isophorone	34	0.30	1.0 ug/L	40.0	ND	85.0	21-196	2.21	93	
2-Methylnaphthalene	35	1.0	2.0 ug/L	40.0	ND	86.8	44-106	4.81	33	
2-Methylphenol (o-cresol)	29	0.60	5.0 ug/L	40.0	ND	73.4	9-124	3.31	33	
3 & 4-Methylphenol (m & p-cresol)	26	0.60	5.0 ug/L	40.0	ND	65.8	12-99	10.8	33	
N-Nitrosodi-n-propylamine	33	0.80	2.0 ug/L	40.0	ND	81.8	2-230	9.39	87	
N-Nitrosodimethylamine	20	0.30	2.0 ug/L	40.0	ND	51.2	10-84	5.69	59	
N-Nitrosodiphenylamine	38	1.0	1.0 ug/L	40.0	ND	94.8	40-115	6.18	25	
Naphthalene	33	0.70	1.0 ug/L	40.0	ND	82.1	21-133	2.59	65	
2-Nitroaniline	58	1.0	20 ug/L	60.0	ND	96.0	48-128	2.45	26	
3-Nitroaniline	59	5.0	50 ug/L	60.0	ND	98.3	17-137	4.33	25	
4-Nitroaniline	61	4.0	20 ug/L	60.0	ND	102	13-191	10.4	33	
2-Nitrophenol	31	0.90	10 ug/L	40.0	ND	77.9	29-182	5.85	55	
4-Nitrophenol	8.0	3.0	5.0 ug/L	40.0	ND	20.0	7-132	4.09	131	
Nitrobenzene	33	0.90	1.0 ug/L	40.0	ND	82.4	35-180	2.90	62	
Pentachlorophenol	ND	4.0	5.0 ug/L	40.0	ND		14-176		86	QM-05, U
Phenanthrene	38	0.90	2.0 ug/L	40.0	ND	95.5	54-120	1.87	39	
Phenol	16	0.50	1.0 ug/L	40.0	ND	40.2	5-120	6.50	64	
Pyrene	37	0.40	2.0 ug/L	40.0	ND	93.0	52-120	4.80	49	
1,2,4-Trichlorobenzene	32	0.60	2.0 ug/L	40.0	ND	80.4	44-142	0.0933	50	
2,4,5-Trichlorophenol	34	0.60	2.0 ug/L	40.0	ND	84.6	65-125	5.91	25	
2,4,6-Trichlorophenol	28	0.70	2.0 ug/L	40.0	ND	69.0	37-144	5.24	58	
Surrogate: 2-Fluorobiphenyl	39.4		ug/L	40.0		98.4	27-119			
Surrogate: 2-Fluorophenol	22.9		ug/L	40.0		57.2	7-85			
Surrogate: Nitrobenzene-d5	38.3		ug/L	40.0		95.7	15-314			
Surrogate: p-Terphenyl-d14	36.7		ug/L	40.0		91.7	36-141			
Surrogate: Phenol-d6	15.0		ug/L	40.0		37.4	1-65			
Surrogate: 2,4,6-Tribromophenol	29.8		ug/L	40.0		74.4	4-168			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Benzidines by EPA Method 8270C - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13243 - SVOAs in Water GCMS

Blank (AJ13243-BLK1)		Prepared: 10/05/21 Analyzed: 10/11/21									
Benzidine	ND	3.0	5.0	ug/L							U
3,3'-Dichlorobenzidine	ND	2.0	5.0	ug/L							U
Surrogate: 2-Fluorobiphenyl	30.4			ug/L	40.0		76.0	27-119			
Surrogate: Nitrobenzene-d5	30.2			ug/L	40.0		75.5	15-314			
Surrogate: p-Terphenyl-d14	32.6			ug/L	40.0		81.6	36-141			

LCS (AJ13243-BS1)		Prepared: 10/05/21 Analyzed: 10/11/21									
Benzidine	13.0	3.0	5.0	ug/L	79.4		16.4	12-151			
3,3'-Dichlorobenzidine	74.6	2.0	5.0	ug/L	79.4		94.0	8-120			
Surrogate: 2-Fluorobiphenyl	31.8			ug/L	40.0		79.4	27-119			
Surrogate: Nitrobenzene-d5	31.4			ug/L	40.0		78.4	15-314			
Surrogate: p-Terphenyl-d14	31.8			ug/L	40.0		79.6	36-141			

Matrix Spike (AJ13243-MS1)		Source: 2113566-01		Prepared: 10/05/21 Analyzed: 10/11/21							
Benzidine	51.0	3.0	5.0	ug/L	79.4	ND	64.3	4-262			
3,3'-Dichlorobenzidine	73.4	2.0	5.0	ug/L	79.4	ND	92.5	4-262			
Surrogate: 2-Fluorobiphenyl	29.4			ug/L	40.0		73.6	27-119			
Surrogate: Nitrobenzene-d5	28.7			ug/L	40.0		71.8	15-314			
Surrogate: p-Terphenyl-d14	33.1			ug/L	40.0		82.8	36-141			

Matrix Spike Dup (AJ13243-MSD1)		Source: 2113566-01		Prepared: 10/05/21 Analyzed: 10/11/21							
Benzidine	61.9	3.0	5.0	ug/L	79.4	ND	78.0	4-262	19.3	108	
3,3'-Dichlorobenzidine	76.2	2.0	5.0	ug/L	79.4	ND	96.0	4-262	3.69	108	
Surrogate: 2-Fluorobiphenyl	30.6			ug/L	40.0		76.5	27-119			
Surrogate: Nitrobenzene-d5	30.9			ug/L	40.0		77.2	15-314			
Surrogate: p-Terphenyl-d14	31.9			ug/L	40.0		79.8	36-141			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C SIM - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13401 - SVOAs in Water GCMS

Blank (AJ13401-BLK1)

Prepared: 10/05/21 Analyzed: 10/07/21

Acenaphthylene	ND	0.20	0.20	ug/L							U
Anthracene	ND	0.090	0.20	ug/L							U
Benzo (a) anthracene	ND	0.20	0.20	ug/L							U
Benzo (a) pyrene	ND	0.20	0.20	ug/L							U
Benzo (b) fluoranthene	ND	0.20	0.20	ug/L							U
Benzo (g,h,i) perylene	ND	0.10	0.20	ug/L							U
Benzo (k) fluoranthene	ND	0.20	0.20	ug/L							U
Chrysene	ND	0.20	0.20	ug/L							U
Dibenz (a,h) anthracene	ND	0.10	0.20	ug/L							U
Fluorene	ND	0.20	0.20	ug/L							U
Indeno (1,2,3-cd) pyrene	ND	0.030	0.20	ug/L							U
Phenanthrene	ND	0.10	0.20	ug/L							U
Pyrene	ND	0.20	0.20	ug/L							U
Surrogate: 2-Fluorobiphenyl	32.2			ug/L	40.0		80.5	24-119			
Surrogate: Nitrobenzene-d5	35.4			ug/L	40.0		88.5	15-314			
Surrogate: p-Terphenyl-d14	40.8			ug/L	40.0		102	37-139			

LCS (AJ13401-BS1)

Prepared: 10/05/21 Analyzed: 10/07/21

Acenaphthylene	38.4	0.20	0.20	ug/L	40.0		96.0	54-126			
Anthracene	29.8	0.090	0.20	ug/L	40.0		74.4	43-120			
Benzo (a) anthracene	38.2	0.20	0.20	ug/L	40.0		95.4	42-133			
Benzo (a) pyrene	43.7	0.20	0.20	ug/L	40.0		109	32-148			
Benzo (b) fluoranthene	42.6	0.20	0.20	ug/L	40.0		106	42-140			
Benzo (g,h,i) perylene	42.9	0.10	0.20	ug/L	40.0		107	2-195			
Benzo (k) fluoranthene	51.2	0.20	0.20	ug/L	40.0		128	25-146			
Chrysene	38.7	0.20	0.20	ug/L	40.0		96.8	44-140			
Dibenz (a,h) anthracene	44.2	0.10	0.20	ug/L	40.0		111	2-200			
Fluorene	41.1	0.20	0.20	ug/L	40.0		103	70-120			
Indeno (1,2,3-cd) pyrene	44.5	0.030	0.20	ug/L	40.0		111	2-151			
Phenanthrene	35.4	0.10	0.20	ug/L	40.0		88.4	65-120			
Pyrene	39.3	0.20	0.20	ug/L	40.0		98.2	70-120			
Surrogate: 2-Fluorobiphenyl	36.5			ug/L	40.0		91.3	24-119			
Surrogate: Nitrobenzene-d5	37.7			ug/L	40.0		94.2	15-314			
Surrogate: p-Terphenyl-d14	39.9			ug/L	40.0		99.8	37-139			

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
 Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC 262 Rickenbacker Circle Livermore CA, 94551	Project Manager: Quality Control Manager Project: QC- 32oz Glass Boston Round Amber Project Number: Lot Number G1271CUFS	Reported: 10/27/21 11:37
---	--	-----------------------------

Semivolatile Organic Compounds by EPA Method 8270C SIM - Quality Control

Analyte	Result	MDL	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch AJ13401 - SVOAs in Water GCMS

Matrix Spike (AJ13401-MS1)		Source: 2113566-01			Prepared: 10/05/21		Analyzed: 10/07/21	
Acenaphthylene	36.9	0.20	0.20	ug/L	40.0	ND	92.3	33-145
Anthracene	34.6	0.090	0.20	ug/L	40.0	ND	86.4	27-133
Benzo (a) anthracene	33.3	0.20	0.20	ug/L	40.0	ND	83.3	33-143
Benzo (a) pyrene	25.2	0.20	0.20	ug/L	40.0	ND	63.1	17-163
Benzo (b) fluoranthene	26.8	0.20	0.20	ug/L	40.0	ND	66.9	24-159
Benzo (g,h,i) perylene	16.4	0.10	0.20	ug/L	40.0	ND	40.9	2-219
Benzo (k) fluoranthene	28.6	0.20	0.20	ug/L	40.0	ND	71.6	11-162
Chrysene	32.8	0.20	0.20	ug/L	40.0	ND	82.0	17-168
Dibenz (a,h) anthracene	16.8	0.10	0.20	ug/L	40.0	ND	42.0	2-227
Fluorene	37.6	0.20	0.20	ug/L	40.0	ND	94.1	59-121
Indeno (1,2,3-cd) pyrene	17.9	0.030	0.20	ug/L	40.0	ND	44.8	2-171
Phenanthrene	33.5	0.10	0.20	ug/L	40.0	ND	83.8	54-120
Pyrene	35.0	0.20	0.20	ug/L	40.0	ND	87.6	52-120
Surrogate: 2-Fluorobiphenyl	33.3			ug/L	40.0		83.2	24-119
Surrogate: Nitrobenzene-d5	35.4			ug/L	40.0		88.4	15-314
Surrogate: p-Terphenyl-d14	31.0			ug/L	40.0		77.6	37-139

Matrix Spike Dup (AJ13401-MSD1)		Source: 2113566-01			Prepared: 10/05/21		Analyzed: 10/07/21			
Acenaphthylene	34.1	0.20	0.20	ug/L	40.0	ND	85.2	33-145	8.03	74
Anthracene	29.7	0.090	0.20	ug/L	40.0	ND	74.2	27-133	15.3	66
Benzo (a) anthracene	31.3	0.20	0.20	ug/L	40.0	ND	78.2	33-143	6.38	53
Benzo (a) pyrene	26.1	0.20	0.20	ug/L	40.0	ND	65.4	17-163	3.50	72
Benzo (b) fluoranthene	30.2	0.20	0.20	ug/L	40.0	ND	75.4	24-159	12.0	71
Benzo (g,h,i) perylene	17.4	0.10	0.20	ug/L	40.0	ND	43.4	2-219	5.81	97
Benzo (k) fluoranthene	24.8	0.20	0.20	ug/L	40.0	ND	62.0	11-162	14.3	63
Chrysene	34.3	0.20	0.20	ug/L	40.0	ND	85.7	17-168	4.50	87
Dibenz (a,h) anthracene	18.6	0.10	0.20	ug/L	40.0	ND	46.5	2-227	10.3	126
Fluorene	39.1	0.20	0.20	ug/L	40.0	ND	97.8	59-121	3.91	38
Indeno (1,2,3-cd) pyrene	17.6	0.030	0.20	ug/L	40.0	ND	44.1	2-171	1.57	99
Phenanthrene	33.2	0.10	0.20	ug/L	40.0	ND	82.9	54-120	1.05	39
Pyrene	35.4	0.20	0.20	ug/L	40.0	ND	88.5	52-120	0.994	49
Surrogate: 2-Fluorobiphenyl	39.2			ug/L	40.0		98.0	24-119		
Surrogate: Nitrobenzene-d5	45.5			ug/L	40.0		114	15-314		
Surrogate: p-Terphenyl-d14	37.5			ug/L	40.0		93.8	37-139		

The results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



Alpha Analytical Laboratories, Inc. email: clientservices@alpha-labs.com
Corporate: 208 Mason Street | Ukiah, CA 95482 | T: 707-468-0401 | F: 707-468-5267 | ELAP# 1551

Sample Traps, LLC
262 Rickenbacker Circle
Livermore CA, 94551

Project Manager: Quality Control Manager
Project: QC- 32oz Glass Boston Round Amber
Project Number: Lot Number G1271CUFS

Reported:
10/27/21 11:37

Notes and Definitions

- J Detected but below the Reporting Limit; therefore, result is an estimated concentration, detected but not quantified (DNQ).
- QM-05 The spike recovery was outside acceptance limits for the MS and/or MSD due to matrix interference. The LCS and/or LCSD were within acceptance limits showing that the laboratory is in control and the data is acceptable.
- T-06 Sample analyzed outside of recommended holding time.
- U Analyte included in analysis, but not detected at or above MDL.
- ND Analyte NOT DETECTED at or above the reporting limit
- dry Sample results reported on a dry weight basis
- MDL Method detection limit
- Rec Recovery
- RPD Relative Percent Difference

Non-accredited analytes are reported only when ELAP accreditation for a requested analyte is not available. For a list of accredited analytes, view our certificates at the Company link on our website at www.alpha-labs.com or contact your Project Manager directly.



Laboratory & Corporate: 208 Mason Street, Ukiah, CA 95482
707-468-0401 Fax: 707-468-5267

Service Center & Micro Lab: 6398 Dougherty Rd, Ste 35, Dublin, CA 94568
925-828-6226 Fax: 925-828-6309

Chain of Custody Record

Reports and Invoices will be delivered by email in .pdf format.

Lab No. 2150119 Page of

Ukiah 30

Report to:		Invoice to (if different):		Project Info for Report:		Signature below authorizes work under terms stated on reverse side.																	
Company:		Company:		Project ID:		Analyses Requested					TAT	Sample Notes (lab use only)											
Sample Traps LLC				QC- 32oz Glass Boston Round Amber																			
Attn:		Attn:		Project No:		: Total Number of Containers	include j-flag reporting	608 Sample Traps	8141 Sample Traps	8151 Sample Traps	8270 Sample Traps Group	8270 PNA SIM Sample Traps	10 days	Sample Notes									
Quality Control Manager				Lot number G1271CUFS											5 days	Temperature:							
Address:		Address:		PO/Reference :		48 hours	Other:	deg. C	Shipment Method:														
Phone/Fax:		Phone/Fax:				days	Lab Approval Required For Rush TATs	Custody Seals:	Y / N														
Email Address:		Email Address:				Sample Notes or CDPH Source Numbers:	Y	N															
admin@sampletraps.com																							
Samplers Signature:		Container:		Preservative:		Matrix:																	
Print:		40ml VOA		Poly		Glass bottle		Glass Jar		HCL		Methanol		Na Bisulfate		Other		None		Water		Container	
Sample Identification		Sampled: Date Time																					
G1271CUFS - 01																							
G1271CUFS - 02																							
G1271CUFS - 03																							
G1271CUFS - 04																							
G1271CUFS - 05																							
Relinquished by:		Received by:		Date:		Time:		CDPH Write On EDT Transmission? <input type="radio"/> Yes <input type="radio"/> No															
<i>[Signature]</i>		<i>JE</i>		<i>9/28/21</i>		<i>1940</i>		State System Number: _____															
		<i>JE</i>		<i>9-23-21</i>		<i>232</i>		If "Y" please enter the Source Number(s) in the column above															
								CA Geotracker EDF Report? <input type="radio"/> Yes <input type="radio"/> No															
								Global ID: _____ Sampling Company Log Code: _____															
								EDF to (Email Address): _____															
								Travel and Site Time: _____ Mileage: _____ Misc. Supplies: _____															